

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436
 Lab File ID: VY022872.D Date Analyzed: 06/30/2025
 Instrument ID: MSVOA_Y Time Analyzed: 11:03
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	450935	7.71	753538	8.62	651538	11.41
UPPER LIMIT	901870	8.207	1507080	9.115	1303080	11.914
LOWER LIMIT	225468	7.207	376769	8.115	325769	10.914
EPA SAMPLE NO.						
TP-85	372972	7.71	690751	8.61	657778	11.41
TP-86	346216	7.71	639767	8.62	633746	11.41
TP-84	323652	7.71	612070	8.61	599073	11.41
VY0630SBL01	371114	7.71	672085	8.62	635637	11.41
VY0630SBS01	473861	7.71	796065	8.62	673720	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.