

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436
 Lab File ID: VY022852.D Date Analyzed: 06/27/2025
 Instrument ID: MSVOA_Y Time Analyzed: 11:00
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	319930	13.347				
UPPER LIMIT	639860	13.847				
LOWER LIMIT	159965	12.847				
EPA SAMPLE NO.						
TP-70	238564	13.35				
TP-69	255220	13.35				
TP-83	232255	13.35				
TP-87	204342	13.35				
TP-100	240211	13.35				
TP-99	230605	13.35				
TP-82	250591	13.35				
VY0627SBL01	250491	13.35				
VY0627SBS01	290353	13.35				
VY0627SBSD01	274040	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.