

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436
 Lab File ID: VY022872.D Date Analyzed: 06/30/2025
 Instrument ID: MSVOA_Y Time Analyzed: 11:03
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	313014	13.346				
UPPER LIMIT	626028	13.846				
LOWER LIMIT	156507	12.846				
EPA SAMPLE NO.						
TP-85	284435	13.34				
TP-86	290240	13.35				
TP-84	250413	13.34				
VY0630SBL01	274914	13.35				
VY0630SBS01	318118	13.34				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.