

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CHEM02

Lab Code: ACE

SDG NO.: Q2437

Continuing Calib Date: 07/02/2025

Initial Calibration Date(s): 06/09/2025

06/09/2025

Continuing Calib Time: 09:29

Initial Calibration Time(s): 09:14

18:44

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.48	4.49	4.42	4.56	0.01
Aroclor-1016-2	(2)	4.50	4.50	4.43	4.57	0.00
Aroclor-1016-3	(3)	4.56	4.56	4.49	4.63	0.00
Aroclor-1016-4	(4)	4.65	4.65	4.58	4.72	0.00
Aroclor-1016-5	(5)	4.93	4.94	4.87	5.01	0.01
Aroclor-1260-1	(1)	6.02	6.03	5.96	6.10	0.01
Aroclor-1260-2	(2)	6.28	6.28	6.21	6.35	0.00
Aroclor-1260-3	(3)	6.63	6.63	6.56	6.70	0.00
Aroclor-1260-4	(4)	6.85	6.85	6.78	6.92	0.01
Aroclor-1260-5	(5)	7.15	7.15	7.08	7.22	0.00
TCX		3.40	3.40	3.35	3.45	0.00
DCB		8.50	8.50	8.40	8.60	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CHEM02

Lab Code: ACE

SDG NO.: Q2437

Continuing Calib Date: 07/02/2025

Initial Calibration Date(s): 06/09/2025 06/09/2025

Continuing Calib Time: 09:29

Initial Calibration Time(s): 09:14 18:44

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	3.78	3.78	3.71	3.85	0.00
Aroclor-1016-2	(2)	3.80	3.80	3.73	3.87	0.00
Aroclor-1016-3	(3)	3.96	3.96	3.89	4.03	0.00
Aroclor-1016-4	(4)	4.01	4.01	3.94	4.08	0.00
Aroclor-1016-5	(5)	4.20	4.20	4.13	4.27	0.00
Aroclor-1260-1	(1)	5.19	5.20	5.13	5.27	0.01
Aroclor-1260-2	(2)	5.38	5.39	5.32	5.46	0.01
Aroclor-1260-3	(3)	5.52	5.53	5.46	5.60	0.01
Aroclor-1260-4	(4)	5.98	5.99	5.92	6.06	0.01
Aroclor-1260-5	(5)	6.23	6.23	6.16	6.30	0.00
TCX		2.78	2.78	2.73	2.83	0.00
DCB		7.54	7.54	7.44	7.64	0.00