

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2437 SAS No.: Q2437 SDG NO.: Q2437

EPA Sample No.: SSTD020674 Date Analyzed: 07/01/2025

Lab File ID: BP025052.D Time Analyzed: 11:57

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	316339	7.443	1330390	10.19	879629	14.10
UPPER LIMIT	632678	7.943	2660780	10.69	1759260	14.595
LOWER LIMIT	158170	6.943	665195	9.69	439815	13.595
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	332832	7.44	1299900	10.19	829789	14.10

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.