

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2437 SAS No.: Q2437 SDG NO.: Q2437

EPA Sample No.: SSTD020674 Date Analyzed: 07/01/2025

Lab File ID: BP025052.D Time Analyzed: 11:57

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1819960	16.913	2010050	21.348	2174320	24.477
UPPER LIMIT	3639920	17.413	4020100	21.848	4348640	24.977
LOWER LIMIT	909980	16.413	1005030	20.848	1087160	23.977
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	1760750	16.92	1847010	21.35	1934230	24.50

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.