

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CHEM02

Lab Code: CHEM Case No.: Q2437 SAS No.: Q2437 SDG No.: Q2437

Instrument ID: MSVOA_N Calibration Date/Time: 06/27/2025 09:38

Lab File ID: VN087207.D Init. Calib. Date(s): 06/26/2025 06/26/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 16:34 18:39

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Ethyl Acetate	50.000	46.25	7.498001	20
1,1,2-Trichlorotrifluoroethane	50.000	45.08	9.835999	20
Acrylonitrile	250.000	236.11	5.556	20
Acetone	250.000	285.25	-14.10081	20
Carbon Disulfide	50.000	50.80	-1.601997	20
Methyl Acetate	50.000	43.90	12.194	20
Methylene Chloride	50.000	48.96	2.082001	20
4-Methyl-2-Pentanone	250.000	227.29	9.085199	20
2-Chloroethyl Vinyl ether	250.000	226.39	9.444397	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.