

Surrogate Summary

SDG No.: Q2437

Client: Chemtech Consulting Group

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2437-01	STORAGE-BLANK-SOIL-REF-6	1,2-Dichloroethane-d4	50	61.4	123	74	125
		Dibromofluoromethane	50	46.9	94	75	124
		Toluene-d8	50	46.2	92	86	113
		4-Bromofluorobenzene	50	46.3	93	77	121
Q2437-02	STORAGE-BLANK-WATER-REF-5	1,2-Dichloroethane-d4	50	56.0	112	74	125
		Dibromofluoromethane	50	38.5	77	75	124
		Toluene-d8	50	45.8	92	86	113
		4-Bromofluorobenzene	50	39.8	80	77	121
Q2437-03	STORAGE-BLANK-WATER-REF-4	1,2-Dichloroethane-d4	50	58.4	117	74	125
		Dibromofluoromethane	50	51.3	103	75	124
		Toluene-d8	50	47.0	94	86	113
		4-Bromofluorobenzene	50	44.7	89	77	121
Q2437-04	STORAGE-BLANK-SAMPLE-MANAC	1,2-Dichloroethane-d4	50	60.4	121	74	125
		Dibromofluoromethane	50	46.9	94	75	124
		Toluene-d8	50	45.6	91	86	113
		4-Bromofluorobenzene	50	44.8	90	77	121
VN0627WBL01	VN0627WBL01	1,2-Dichloroethane-d4	50	56.5	113	74	125
		Dibromofluoromethane	50	48.6	97	75	124
		Toluene-d8	50	45.1	90	86	113
		4-Bromofluorobenzene	50	42.7	85	77	121
VN0627WBS01	VN0627WBS01	1,2-Dichloroethane-d4	50	51.2	102	74	125
		Dibromofluoromethane	50	48.5	97	75	124
		Toluene-d8	50	48.4	97	86	113
		4-Bromofluorobenzene	50	50.9	102	77	121
VN0627WBSD0	VN0627WBSD01	1,2-Dichloroethane-d4	50	49.1	98	74	125
		Dibromofluoromethane	50	56.8	114	75	124
		Toluene-d8	50	48.0	96	86	113
		4-Bromofluorobenzene	50	49.7	99	77	121

