

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: Q2437 SAS No.: Q2437 SDG NO.: Q2437
 Lab File ID: VN087207.D Date Analyzed: 06/27/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:38
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	196333	8.23	346797	9.11	266770	11.87
UPPER LIMIT	392666	8.73	693594	9.606	533540	12.365
LOWER LIMIT	98166.5	7.73	173399	8.606	133385	11.365
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	132710	8.23	292238	9.11	265444	11.87
STORAGE-BLANK-WATER-REF-5	184837	8.23	453762	9.11	401647	11.87
STORAGE-BLANK-WATER-REF-4	138288	8.23	271970	9.11	247376	11.87
STORAGE-BLANK-SAMPLE-MANAGEMENT	136871	8.23	285243	9.11	250172	11.87
VN0627WBL01	196165	8.23	398772	9.10	333801	11.87
VN0627WBS01	176379	8.23	277531	9.11	254005	11.87
VN0627WBSD01	143301	8.23	249904	9.11	238840	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.