

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: Q2437 SAS No.: Q2437 SDG NO.: Q2437
 Lab File ID: VN087207.D Date Analyzed: 06/27/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:38
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	145043	13.788				
UPPER LIMIT	290086	14.288				
LOWER LIMIT	72521.5	13.288				
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	120316	13.79				
STORAGE-BLANK-WATER-REF-5	188552	13.79				
STORAGE-BLANK-WATER-REF-4	113772	13.79				
STORAGE-BLANK-SAMPLE-MANAGEMENT	113401	13.79				
VN0627WBL01	147923	13.79				
VN0627WBS01	127421	13.79				
VN0627WBSD01	118817	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.