



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Tetra Tech NUS, Inc.

Project Name: NWIRP Bethpage 112G08005-WE13

Project Manager : Ernie Wu

Order ID # Q2439

Test Name: TDS,TSS

A. Number of Samples and Date of Receipt:

2 Water samples were received on 06/27/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Metals Group4, SVOC-SIMGroup1, TDS and TSS. This data package contains results for TDS,TSS.

C. Analytical Techniques:

The analysis of TDS was based on method SM2540 C and The analysis of TSS was based on method SM2540 D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all parameters.

The Duplicate analysis met criteria for all parameters.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is).

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____