

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136345

Review By	Janvi	Review On	7/2/2025 4:45:37 PM
Supervise By	jaswal	Supervise On	7/2/2025 11:18:53 PM

STD. NAME	STD REF.#
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212
ICV Standard	MP86213
CCV Standard	MP86216
ICSA Standard	MP86214,MP86215
CRI Standard	MP86212
LCS Standard	
Chk Standard	MP86217,MP86218

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	07/01/25 15:28		Janvi	OK
2	S1	S1	CAL2	07/01/25 15:33		Janvi	OK
3	S2	S2	CAL3	07/01/25 15:37		Janvi	OK
4	S3	S3	CAL4	07/01/25 15:42		Janvi	OK
5	S4	S4	CAL5	07/01/25 15:46		Janvi	OK
6	S5	S5	CAL6	07/01/25 15:50		Janvi	OK
7	ICV01	ICV01	ICV	07/01/25 16:17		Janvi	OK
8	LLICV01	LLICV01	LLICV	07/01/25 16:31		Janvi	OK
9	ICB01	ICB01	ICB	07/01/25 16:37		Janvi	OK
10	CRI01	CRI01	CRDL	07/01/25 16:41		Janvi	OK
11	ICSA01	ICSA01	ICSA	07/01/25 16:50		Janvi	OK
12	ICSAB01	ICSAB01	ICSAB	07/01/25 16:54		Janvi	OK
13	ICSADL	ICSADL	ICSA	07/01/25 16:59		Janvi	OK
14	ICSABDL	ICSABDL	ICSAB	07/01/25 17:03		Janvi	OK
15	CCV01	CCV01	CCV	07/01/25 17:07		Janvi	OK
16	CCB01	CCB01	CCB	07/01/25 17:12		Janvi	OK
17	PB168664BL	PB168664BL	MB	07/01/25 17:49		Janvi	OK
18	PB168664BS	PB168664BS	LCS	07/01/25 17:53		Janvi	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136345

Review By	Janvi	Review On	7/2/2025 4:45:37 PM
Supervise By	jaswal	Supervise On	7/2/2025 11:18:53 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86213		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86215		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

19	Q2444-01	1ST-PL-CHILD-CARE	SAM	07/01/25 18:06		Janvi	OK
20	PB168661BS	PB168661BS	LCS	07/01/25 18:11		Janvi	OK
21	Q2444-02	1ST-PL-CHILD-CARE	SAM	07/01/25 18:15		Janvi	OK
22	Q2444-03	1ST-PL-CHILD-CARE	SAM	07/01/25 18:19		Janvi	OK
23	Q2444-04	1ST-PL-CHILD-CARE	SAM	07/01/25 18:24		Janvi	OK
24	Q2444-05	1ST-PL-CHILD-CARE	SAM	07/01/25 18:28		Janvi	OK
25	Q2444-06	1ST-PL-CHILD-CARE	SAM	07/01/25 18:32		Janvi	OK
26	CCV02	CCV02	CCV	07/01/25 18:37		Janvi	OK
27	CCB02	CCB02	CCB	07/01/25 18:41		Janvi	OK
28	Q2444-07	1ST-PL-KITCHENSIN	SAM	07/01/25 18:45		Janvi	OK
29	Q2444-08	1ST-PL-KITCHENSIN	SAM	07/01/25 18:49	Bad Injection	Janvi	Not Ok
30	Q2444-09	1ST-PL-RESTROOM	SAM	07/01/25 18:54		Janvi	OK
31	Q2444-10	1ST-PL-RESTROOM	SAM	07/01/25 18:58		Janvi	OK
32	Q2444-11	1ST-PL-RESTROOM	SAM	07/01/25 19:02		Janvi	OK
33	Q2444-12	1ST-PL-RESTROOM	SAM	07/01/25 19:07		Janvi	OK
34	Q2444-12DUP	1ST-PL-RESTROOM	DUP	07/01/25 19:16		Janvi	OK
35	Q2444-12L	1ST-PL-RESTROOM	SD	07/01/25 19:20		Janvi	OK
36	Q2444-12MS	1ST-PL-RESTROOM	MS	07/01/25 19:24		Janvi	OK
37	Q2444-12MSD	1ST-PL-RESTROOM	MSD	07/01/25 19:29		Janvi	OK
38	CCV03	CCV03	CCV	07/01/25 19:33		Janvi	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136345

Review By	Janvi	Review On	7/2/2025 4:45:37 PM
Supervise By	jaswal	Supervise On	7/2/2025 11:18:53 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86213		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86215		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

39	CCB03	CCB03	CCB	07/01/25 19:37		Janvi	OK
40	Q2444-12A	1ST-PL-RESTROOM	PS	07/01/25 19:42	0.1ml of M6004 and 0.1ml of M6013 in 10ml sample	Janvi	OK
41	CCV04	CCV04	CCV	07/01/25 20:43		Janvi	OK
42	CCB04	CCB04	CCB	07/01/25 20:47		Janvi	OK
43	Q2439-01	RW8-SP100-2025062	SAM	07/01/25 20:56		Janvi	OK
44	Q2439-02	RW8-SP303-2025062	SAM	07/01/25 21:00		Janvi	OK
45	Q2439-02DUP	RW8-SP303-2025062	DUP	07/01/25 21:05		Janvi	OK
46	Q2439-02L	RW8-SP303-2025062	SD	07/01/25 21:09		Janvi	OK
47	Q2439-02MS	RW8-SP303-2025062	MS	07/01/25 21:13		Janvi	OK
48	Q2439-02MSD	RW8-SP303-2025062	MSD	07/01/25 21:18		Janvi	OK
49	Q2439-02A	RW8-SP303-2025062	PS	07/01/25 21:22	0.1ml of M6004 and 0.1ml of M6013 in 10ml sample	Janvi	OK
50	CCV05	CCV05	CCV	07/01/25 21:26		Janvi	OK
51	CCB05	CCB05	CCB	07/01/25 21:31		Janvi	OK
52	PB168661BL	PB168661BL	MB	07/01/25 21:41		Janvi	OK
53	CCV06	CCV06	CCV	07/01/25 21:45		Janvi	OK
54	CCB06	CCB06	CCB	07/01/25 21:49		Janvi	OK