

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q2439

Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q2439 **SAS No.:** Q2439

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	3700	4000	93	90 - 110	P	07/01/2025	16:17	LB136345

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q2439

Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q2439 **SAS No.:** Q2439

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	107	100	107	80 - 120	P	07/01/2025	16:31	LB136345

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q2439

Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q2439 **SAS No.:** Q2439

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	5170	5000	103	90 - 110	P	07/01/2025	17:07	LB136345
CCV02	Iron	4920	5000	98	90 - 110	P	07/01/2025	18:37	LB136345
CCV03	Iron	4790	5000	96	90 - 110	P	07/01/2025	19:33	LB136345
CCV04	Iron	4960	5000	99	90 - 110	P	07/01/2025	20:43	LB136345
CCV05	Iron	4750	5000	95	90 - 110	P	07/01/2025	21:26	LB136345
CCV06	Iron	4770	5000	96	90 - 110	P	07/01/2025	21:45	LB136345