

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tetra Tech NUS, Inc. **SDG No.:** Q2439
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q2439 **SAS No.:** Q2439
ICS Source: EPA **Instrument ID:** P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Iron	91600	100000	92	85600	116500	07/01/2025	16:50	LB136345
ICSAB01	Iron	94600	99000	96	84400	114500	07/01/2025	16:54	LB136345