

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Tetra Tech NUS, Inc. **level:** low **sdg no.:** Q2439
contract: TETR06 **lab code:** CHEM **case no.:** Q2439 **sas no.:** Q2439
matrix: Water **sample id:** Q2439-02 **client id:** RW8-SP303-20250626MS
Percent Solids for Sample: NA **Spiked ID:** Q2439-02MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	87 - 115	1690		89.4		1500	107		P

metals

- 5a -

MATRIX SPIKE DUPLICATE SUMMARY

client: Tetra Tech NUS, Inc. level: low sdg no.: Q2439
contract: TETR06 lab code: CHEM case no.: Q2439 sas no.: Q2439
matrix: Water sample id: Q2439-02 client id: RW8-SP303-20250626MSD
Percent Solids for Sample: NA Spiked ID: Q2439-02MSD Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	87 - 115	1560		89.4		1500	98		P