

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2439 SAS No.: Q2439 SDG NO.: Q2439

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 07/02/2025

Lab File ID: BN037425.D Time Analyzed: 14:04

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2413	7.561	5075	10.34	3293	14.21
UPPER LIMIT	4826	8.061	10150	10.84	6586	14.713
LOWER LIMIT	1206.5	7.061	2537.5	9.84	1646.5	13.713
EPA SAMPLE NO.						
01 PB168695BL	2320	7.56	4227	10.34	2645	14.21
02 RW8-SP100-20250626	1716	7.56	3491	10.34	2495	14.21
03 PB168695BS	2121	7.56	4355	10.34	2709	14.21
04 PB168695BSD	1957	7.56	4090	10.34	2586	14.21
05 RW8-SP303-20250626	1779	7.57	3739	10.34	2590	14.21

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.