

LAB CHRONICLE

OrderID:	Q2441	OrderDate:	6/27/2025 11:14:45 AM
Client:	Europastry	Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ
Contact:	Kevin Carlucci	Location:	D41

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2441-01	MH-6272025	WATER			06/27/25 09:00			06/27/25
			Oil and Grease	1664A			07/02/25 09:00	
			TPH	1664A			07/02/25 10:00	



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	Europastry	Date Collected:	06/27/25 09:00
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Date Received:	06/27/25
Client Sample ID:	MH-6272025	SDG No.:	Q2441
Lab Sample ID:	Q2441-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Oil and Grease	648		1	0.29	5.00	mg/L		07/02/25 09:00	1664A
TPH	378		1	0.29	5.00	mg/L		07/02/25 10:00	1664A

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

QC RESULT SUMMARY

Preparation Blank Summary

Client: Europastry

SDG No.: Q2441

Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB136347BL							
Oil and Grease	mg/L	< 2.5000	2.5000	U	0.29	5.0	07/02/2025
Sample ID: LB136348BL							
TPH	mg/L	< 2.5000	2.5000	U	0.29	5.0	07/02/2025

Matrix Spike Summary

Client:	Europastry	SDG No.:	Q2441
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q2441-01
Client ID:	MH-6272025MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	668		648		20.0	1	100		07/02/2025

Matrix Spike Summary

Client:	Europastry	SDG No.:	Q2441
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q2441-01
Client ID:	MH-6272025MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	668		648		20.0	1	101		07/02/2025

Duplicate Sample Summary

Client:	Europastray	SDG No.:	Q2441
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	LB136348BS
Client ID:	LB136348BSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TPH	mg/L	+/-18	17.0		17.0		1	0		07/02/2025

Duplicate Sample Summary

Client:	Europastry	SDG No.:	Q2441
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q2441-01
Client ID:	MH-6272025MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	668		668		1	0.03		07/02/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q2441
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB136347

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136347BS							
Oil and Grease	mg/L	20.0	16.9		84	1	78-114	07/02/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q2441
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB136348

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136348BS							
TPH	mg/L	20.0	17.0		85	1	78-114	07/02/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q2441
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB136348

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136348BSD							
TPH	mg/L	20.0	17.0		85	1	78-114	07/02/2025