

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc. **SDG No.:** Q2444
Contract: TOTA02 **Lab Code:** CHEM **Case No.:** Q2444 **SAS No.:** Q2444
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	3860	4000	96	95 - 105	P	07/01/2025	16:17	LB136345

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc. **SDG No.:** Q2444
Contract: TOTA02 **Lab Code:** CHEM **Case No.:** Q2444 **SAS No.:** Q2444
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	11.1	12.0	92	80 - 120	P	07/01/2025	16:31	LB136345

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc. **SDG No.:** Q2444
Contract: TOTA02 **Lab Code:** CHEM **Case No.:** Q2444 **SAS No.:** Q2444
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4860	5000	97	90 - 110	P	07/01/2025	17:07	LB136345
CCV02	Lead	4810	5000	96	90 - 110	P	07/01/2025	18:37	LB136345
CCV03	Lead	4760	5000	95	90 - 110	P	07/01/2025	19:33	LB136345
CCV04	Lead	4890	5000	98	90 - 110	P	07/01/2025	20:43	LB136345
CCV05	Lead	4870	5000	97	90 - 110	P	07/01/2025	21:26	LB136345
CCV06	Lead	4810	5000	96	90 - 110	P	07/01/2025	21:45	LB136345

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc. **SDG No.:** Q2444
Contract: TOTA02 **Lab Code:** CHEM **Case No.:** Q2444 **SAS No.:** Q2444
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	3960	4000	99	95 - 105	P	07/03/2025	14:50	LB136377

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc. **SDG No.:** Q2444
Contract: TOTA02 **Lab Code:** CHEM **Case No.:** Q2444 **SAS No.:** Q2444
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	10.9	12.0	91	80 - 120	P	07/03/2025	14:55	LB136377

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc. **SDG No.:** Q2444
Contract: TOTA02 **Lab Code:** CHEM **Case No.:** Q2444 **SAS No.:** Q2444
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4690	5000	94	90 - 110	P	07/03/2025	16:06	LB136377
CCV02	Lead	4740	5000	95	90 - 110	P	07/03/2025	16:53	LB136377
CCV03	Lead	4720	5000	94	90 - 110	P	07/03/2025	17:28	LB136377