



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Total Quality Environmental Inc.

SDG No.: Q2444

Contract: TOTA02

Lab Code: CHEM

Case No.: Q2444

SAS No.: Q2444

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Lead	9.47	12.0	79	65 - 135	P	07/01/2025	16:41	LB136345
CRI01	Lead	11.1	12.0	92	65 - 135	P	07/03/2025	15:37	LB136377