

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Total Quality Environmental Inc. **SDG No.:** Q2444
Contract: TOTA02 **Lab Code:** CHEM **Case No.:** Q2444 **SAS No.:** Q2444
ICS Source: EPA **Instrument ID:** P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Lead	-7.21			-12	12	07/01/2025	16:50	LB136345
ICSAB01	Lead	40.7	49.0	83	37	61	07/01/2025	16:54	LB136345
ICSA	Lead	12.8			-12	12	07/01/2025	16:59	LB136345
ICSAB	Lead	35.0	49.0	71	37	61	07/01/2025	17:03	LB136345
ICSA01	Lead	1.07			-12	12	07/03/2025	15:41	LB136377
ICSAB01	Lead	44.5	49.0	91	37	61	07/03/2025	15:45	LB136377
ICSA	Lead	53.6			-12	12	07/03/2025	15:50	LB136377
ICSAB	Lead	87.4	49.0	178	37	61	07/03/2025	15:54	LB136377