

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Total Quality Environmental Inc. **level:** low **sdg no.:** Q2444
contract: TOTA02 **lab code:** CHEM **case no.:** Q2444 **sas no.:** Q2444
matrix: Water **sample id:** Q2444-12 **client id:** 1ST-PL-RESTROOM-3-CHILD-CARE-SINK-2MS
Percent Solids for Sample: NA **Spiked ID:** Q2444-12MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Lead	ug/L	75 - 125	489		1.50	J	500	98		P

metals

- 5a -

MATRIX SPIKE DUPLICATE SUMMARY

client: Total Quality Environmental Inc. **level:** low **sdg no.:** Q2444
contract: TOTA02 **lab code:** CHEM **case no.:** Q2444 **sas no.:** Q2444
matrix: Water **sample id:** Q2444-12 **client id:** 1ST-PL-RESTROOM-3-CHILD-CARE-SINK-2M
Percent Solids for Sample: NA **Spiked ID:** Q2444-12MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Lead	ug/L	75 - 125	465		1.50	J	500	93		P