

Surrogate Summary

SDG No.: **Q2455**

Client: **M2 Associates**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2455-01	MW1	1,2-Dichloroethane-d4	50	48.3	97		70 (74)	130 (125)
		Dibromofluoromethane	50	47.6	95		70 (75)	130 (124)
		Toluene-d8	50	50.0	100		70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.3	99		70 (77)	130 (121)
Q2455-02	MW3	1,2-Dichloroethane-d4	50	47.6	95		70 (74)	130 (125)
		Dibromofluoromethane	50	46.5	93		70 (75)	130 (124)
		Toluene-d8	50	49.4	99		70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.1	100		70 (77)	130 (121)
Q2455-03	MW10	1,2-Dichloroethane-d4	50	47.5	95		70 (74)	130 (125)
		Dibromofluoromethane	50	46.1	92		70 (75)	130 (124)
		Toluene-d8	50	49.5	99		70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.4	101		70 (77)	130 (121)
Q2455-04	FIELD-BLANK	1,2-Dichloroethane-d4	50	50.3	101		70 (74)	130 (125)
		Dibromofluoromethane	50	47.8	96		70 (75)	130 (124)
		Toluene-d8	50	49.7	99		70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.9	100		70 (77)	130 (121)
VX0703WBL01	VX0703WBL01	1,2-Dichloroethane-d4	50	50.3	101		70 (74)	130 (125)
		Dibromofluoromethane	50	48.6	97		70 (75)	130 (124)
		Toluene-d8	50	49.6	99		70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.6	101		70 (77)	130 (121)
VX0703WBS01	VX0703WBS01	1,2-Dichloroethane-d4	50	49.9	100		70 (74)	130 (125)
		Dibromofluoromethane	50	49.1	98		70 (75)	130 (124)
		Toluene-d8	50	49.6	99		70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.7	101		70 (77)	130 (121)
VX0703WBSD0	VX0703WBSD01	1,2-Dichloroethane-d4	50	51.3	103		70 (74)	130 (125)
		Dibromofluoromethane	50	49.8	100		70 (75)	130 (124)
		Toluene-d8	50	49.3	99		70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.1	102		70 (77)	130 (121)

