

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: M2AS01  
 Lab Code: ACE SDG NO.: Q2455  
 Lab File ID: VX046869.D Date Analyzed: 07/03/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 08:19  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|----------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD    | 416385        | 5.56  | 663948        | 6.76  | 582621        | 10.05  |
| UPPER LIMIT    | 832770        | 6.056 | 1327900       | 7.263 | 1165240       | 10.549 |
| LOWER LIMIT    | 208193        | 5.056 | 331974        | 6.263 | 291311        | 9.549  |
| EPA SAMPLE NO. |               |       |               |       |               |        |
| MW1            | 370531        | 5.56  | 627398        | 6.77  | 574040        | 10.06  |
| MW3            | 569856        | 5.56  | 963880        | 6.77  | 868363        | 10.06  |
| MW10           | 526667        | 5.56  | 896282        | 6.77  | 818636        | 10.05  |
| FIELD-BLANK    | 419216        | 5.56  | 727439        | 6.77  | 669234        | 10.06  |
| VX0703WBL01    | 435949        | 5.56  | 752530        | 6.76  | 691838        | 10.05  |
| VX0703WBS01    | 380935        | 5.56  | 631809        | 6.77  | 572542        | 10.06  |
| VX0703WBSD01   | 339915        | 5.56  | 568093        | 6.76  | 517645        | 10.05  |

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.