

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: M2AS01
 Lab Code: ACE SDG NO.: Q2455
 Lab File ID: VX046869.D Date Analyzed: 07/03/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:19
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	287854	12.018				
UPPER LIMIT	575708	12.518				
LOWER LIMIT	143927	11.518				
EPA SAMPLE NO.						
MW1	275510	12.02				
MW3	447320	12.02				
MW10	401752	12.02				
FIELD-BLANK	336823	12.02				
VX0703WBL01	351915	12.02				
VX0703WBS01	293826	12.02				
VX0703WBSD01	269874	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.