



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2458

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB13	Mercury	0.076	+/-0.2	U	0.20	CV	07/01/2025	11:45	LB136337

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2458

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB40	Mercury	0.076	+/-0.2	U	0.20	CV	07/01/2025	11:49	LB136337
CCB41	Mercury	0.076	+/-0.2	U	0.20	CV	07/01/2025	12:19	LB136337
CCB42	Mercury	0.076	+/-0.2	U	0.20	CV	07/01/2025	12:49	LB136337
CCB43	Mercury	0.076	+/-0.2	U	0.20	CV	07/01/2025	13:13	LB136337

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2458

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB14	Mercury	0.076	+/-0.2	U	0.20	CV	07/02/2025	10:38	LB136352

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2458

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB44	Mercury	0.076	+/-0.2	U	0.20	CV	07/02/2025	10:49	LB136352
CCB45	Mercury	0.076	+/-0.2	U	0.20	CV	07/02/2025	11:40	LB136352
CCB46	Mercury	0.076	+/-0.2	U	0.20	CV	07/02/2025	12:42	LB136352
CCB47	Mercury	0.076	+/-0.2	U	0.20	CV	07/02/2025	13:23	LB136352

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2458

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	07/03/2025	12:21	LB136371
	Antimony	6.76	+/-25	U	50.0	P	07/03/2025	12:21	LB136371
	Arsenic	5.12	+/-10	U	20.0	P	07/03/2025	12:21	LB136371
	Barium	14.6	+/-50	U	100	P	07/03/2025	12:21	LB136371
	Beryllium	0.56	+/-3	U	6.00	P	07/03/2025	12:21	LB136371
	Cadmium	0.50	+/-3	U	6.00	P	07/03/2025	12:21	LB136371
	Calcium	234	+/-1000	U	2000	P	07/03/2025	12:21	LB136371
	Chromium	2.12	+/-5	U	10.0	P	07/03/2025	12:21	LB136371
	Cobalt	2.26	+/-15	U	30.0	P	07/03/2025	12:21	LB136371
	Copper	4.60	+/-10	U	20.0	P	07/03/2025	12:21	LB136371
	Iron	23.4	+/-50	U	100	P	07/03/2025	12:21	LB136371
	Lead	2.30	+/-6	U	12.0	P	07/03/2025	12:21	LB136371
	Magnesium	244	+/-1000	U	2000	P	07/03/2025	12:21	LB136371
	Manganese	5.94	+/-10	U	20.0	P	07/03/2025	12:21	LB136371
	Nickel	3.06	+/-20	U	40.0	P	07/03/2025	12:21	LB136371
	Potassium	918	+/-1000	U	2000	P	07/03/2025	12:21	LB136371
	Selenium	9.64	+/-10	U	20.0	P	07/03/2025	12:21	LB136371
	Silver	1.62	+/-5	U	10.0	P	07/03/2025	12:21	LB136371
	Sodium	868	+/-1000	U	2000	P	07/03/2025	12:21	LB136371
	Thallium	4.38	+/-20	U	40.0	P	07/03/2025	12:21	LB136371
	Vanadium	6.26	+/-20	U	40.0	P	07/03/2025	12:21	LB136371
	Zinc	16.7	+/-20	U	40.0	P	07/03/2025	12:21	LB136371

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2458

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	07/03/2025	13:27	LB136371
	Antimony	6.76	+/-25	U	50.0	P	07/03/2025	13:27	LB136371
	Arsenic	5.12	+/-10	U	20.0	P	07/03/2025	13:27	LB136371
	Barium	14.6	+/-50	U	100	P	07/03/2025	13:27	LB136371
	Beryllium	0.56	+/-3	U	6.00	P	07/03/2025	13:27	LB136371
	Cadmium	0.50	+/-3	U	6.00	P	07/03/2025	13:27	LB136371
	Calcium	234	+/-1000	U	2000	P	07/03/2025	13:27	LB136371
	Chromium	2.12	+/-5	U	10.0	P	07/03/2025	13:27	LB136371
	Cobalt	2.26	+/-15	U	30.0	P	07/03/2025	13:27	LB136371
	Copper	4.60	+/-10	U	20.0	P	07/03/2025	13:27	LB136371
	Iron	23.4	+/-50	U	100	P	07/03/2025	13:27	LB136371
	Lead	2.30	+/-6	U	12.0	P	07/03/2025	13:27	LB136371
	Magnesium	244	+/-1000	U	2000	P	07/03/2025	13:27	LB136371
	Manganese	5.94	+/-10	U	20.0	P	07/03/2025	13:27	LB136371
	Nickel	3.06	+/-20	U	40.0	P	07/03/2025	13:27	LB136371
	Potassium	918	+/-1000	U	2000	P	07/03/2025	13:27	LB136371
	Selenium	9.64	+/-10	U	20.0	P	07/03/2025	13:27	LB136371
	Silver	1.62	+/-5	U	10.0	P	07/03/2025	13:27	LB136371
	Sodium	868	+/-1000	U	2000	P	07/03/2025	13:27	LB136371
	Thallium	4.38	+/-20	U	40.0	P	07/03/2025	13:27	LB136371
	Vanadium	6.26	+/-20	U	40.0	P	07/03/2025	13:27	LB136371
	Zinc	16.7	+/-20	U	40.0	P	07/03/2025	13:27	LB136371
CCB02	Aluminum	11.3	+/-50	U	100	P	07/03/2025	14:55	LB136371
	Antimony	6.76	+/-25	U	50.0	P	07/03/2025	14:55	LB136371
	Arsenic	5.12	+/-10	U	20.0	P	07/03/2025	14:55	LB136371
	Barium	14.6	+/-50	U	100	P	07/03/2025	14:55	LB136371
	Beryllium	0.56	+/-3	U	6.00	P	07/03/2025	14:55	LB136371
	Cadmium	0.50	+/-3	U	6.00	P	07/03/2025	14:55	LB136371
	Calcium	234	+/-1000	U	2000	P	07/03/2025	14:55	LB136371
	Chromium	2.12	+/-5	U	10.0	P	07/03/2025	14:55	LB136371
	Cobalt	2.26	+/-15	U	30.0	P	07/03/2025	14:55	LB136371
	Copper	4.60	+/-10	U	20.0	P	07/03/2025	14:55	LB136371
	Iron	23.4	+/-50	U	100	P	07/03/2025	14:55	LB136371
	Lead	2.30	+/-6	U	12.0	P	07/03/2025	14:55	LB136371
	Magnesium	244	+/-1000	U	2000	P	07/03/2025	14:55	LB136371
	Manganese	5.94	+/-10	U	20.0	P	07/03/2025	14:55	LB136371
	Nickel	3.06	+/-20	U	40.0	P	07/03/2025	14:55	LB136371
	Potassium	918	+/-1000	U	2000	P	07/03/2025	14:55	LB136371
	Selenium	9.64	+/-10	U	20.0	P	07/03/2025	14:55	LB136371

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2458

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	07/03/2025	14:55	LB136371
	Sodium	868	+/-1000	U	2000	P	07/03/2025	14:55	LB136371
	Thallium	4.38	+/-20	U	40.0	P	07/03/2025	14:55	LB136371
	Vanadium	6.26	+/-20	U	40.0	P	07/03/2025	14:55	LB136371
	Zinc	16.7	+/-20	U	40.0	P	07/03/2025	14:55	LB136371
CCB03	Aluminum	11.3	+/-50	U	100	P	07/03/2025	16:08	LB136371
	Antimony	6.76	+/-25	U	50.0	P	07/03/2025	16:08	LB136371
	Arsenic	5.12	+/-10	U	20.0	P	07/03/2025	16:08	LB136371
	Barium	14.6	+/-50	U	100	P	07/03/2025	16:08	LB136371
	Beryllium	0.56	+/-3	U	6.00	P	07/03/2025	16:08	LB136371
	Cadmium	0.50	+/-3	U	6.00	P	07/03/2025	16:08	LB136371
	Calcium	234	+/-1000	U	2000	P	07/03/2025	16:08	LB136371
	Chromium	2.12	+/-5	U	10.0	P	07/03/2025	16:08	LB136371
	Cobalt	2.26	+/-15	U	30.0	P	07/03/2025	16:08	LB136371
	Copper	4.60	+/-10	U	20.0	P	07/03/2025	16:08	LB136371
	Iron	23.4	+/-50	U	100	P	07/03/2025	16:08	LB136371
	Lead	2.30	+/-6	U	12.0	P	07/03/2025	16:08	LB136371
	Magnesium	244	+/-1000	U	2000	P	07/03/2025	16:08	LB136371
	Manganese	5.94	+/-10	U	20.0	P	07/03/2025	16:08	LB136371
	Nickel	3.06	+/-20	U	40.0	P	07/03/2025	16:08	LB136371
	Potassium	918	+/-1000	U	2000	P	07/03/2025	16:08	LB136371
	Selenium	9.64	+/-10	U	20.0	P	07/03/2025	16:08	LB136371
	Silver	1.62	+/-5	U	10.0	P	07/03/2025	16:08	LB136371
	Sodium	868	+/-1000	U	2000	P	07/03/2025	16:08	LB136371
	Thallium	4.38	+/-20	U	40.0	P	07/03/2025	16:08	LB136371
	Vanadium	6.26	+/-20	U	40.0	P	07/03/2025	16:08	LB136371
	Zinc	16.7	+/-20	U	40.0	P	07/03/2025	16:08	LB136371
CCB04	Aluminum	11.3	+/-50	U	100	P	07/03/2025	17:02	LB136371
	Antimony	6.76	+/-25	U	50.0	P	07/03/2025	17:02	LB136371
	Arsenic	5.12	+/-10	U	20.0	P	07/03/2025	17:02	LB136371
	Barium	14.6	+/-50	U	100	P	07/03/2025	17:02	LB136371
	Beryllium	0.56	+/-3	U	6.00	P	07/03/2025	17:02	LB136371
	Cadmium	0.50	+/-3	U	6.00	P	07/03/2025	17:02	LB136371
	Calcium	234	+/-1000	U	2000	P	07/03/2025	17:02	LB136371
	Chromium	2.12	+/-5	U	10.0	P	07/03/2025	17:02	LB136371
	Cobalt	2.26	+/-15	U	30.0	P	07/03/2025	17:02	LB136371
	Copper	4.60	+/-10	U	20.0	P	07/03/2025	17:02	LB136371
	Iron	23.4	+/-50	U	100	P	07/03/2025	17:02	LB136371
	Lead	2.30	+/-6	U	12.0	P	07/03/2025	17:02	LB136371

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2458

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	07/03/2025	17:02	LB136371
	Manganese	5.94	+/-10	U	20.0	P	07/03/2025	17:02	LB136371
	Nickel	3.06	+/-20	U	40.0	P	07/03/2025	17:02	LB136371
	Potassium	918	+/-1000	U	2000	P	07/03/2025	17:02	LB136371
	Selenium	9.64	+/-10	U	20.0	P	07/03/2025	17:02	LB136371
	Silver	1.62	+/-5	U	10.0	P	07/03/2025	17:02	LB136371
	Sodium	868	+/-1000	U	2000	P	07/03/2025	17:02	LB136371
	Thallium	4.38	+/-20	U	40.0	P	07/03/2025	17:02	LB136371
	Vanadium	6.26	+/-20	U	40.0	P	07/03/2025	17:02	LB136371
	Zinc	16.7	+/-20	U	40.0	P	07/03/2025	17:02	LB136371
CCB05	Aluminum	11.3	+/-50	U	100	P	07/03/2025	17:32	LB136371
	Antimony	6.76	+/-25	U	50.0	P	07/03/2025	17:32	LB136371
	Arsenic	5.12	+/-10	U	20.0	P	07/03/2025	17:32	LB136371
	Barium	14.6	+/-50	U	100	P	07/03/2025	17:32	LB136371
	Beryllium	0.56	+/-3	U	6.00	P	07/03/2025	17:32	LB136371
	Cadmium	0.50	+/-3	U	6.00	P	07/03/2025	17:32	LB136371
	Calcium	234	+/-1000	U	2000	P	07/03/2025	17:32	LB136371
	Chromium	2.12	+/-5	U	10.0	P	07/03/2025	17:32	LB136371
	Cobalt	2.26	+/-15	U	30.0	P	07/03/2025	17:32	LB136371
	Copper	4.60	+/-10	U	20.0	P	07/03/2025	17:32	LB136371
	Iron	23.4	+/-50	U	100	P	07/03/2025	17:32	LB136371
	Lead	2.30	+/-6	U	12.0	P	07/03/2025	17:32	LB136371
	Magnesium	244	+/-1000	U	2000	P	07/03/2025	17:32	LB136371
	Manganese	5.94	+/-10	U	20.0	P	07/03/2025	17:32	LB136371
	Nickel	3.06	+/-20	U	40.0	P	07/03/2025	17:32	LB136371
	Potassium	918	+/-1000	U	2000	P	07/03/2025	17:32	LB136371
	Selenium	9.64	+/-10	U	20.0	P	07/03/2025	17:32	LB136371
	Silver	1.62	+/-5	U	10.0	P	07/03/2025	17:32	LB136371
	Sodium	868	+/-1000	U	2000	P	07/03/2025	17:32	LB136371
	Thallium	4.38	+/-20	U	40.0	P	07/03/2025	17:32	LB136371
	Vanadium	6.26	+/-20	U	40.0	P	07/03/2025	17:32	LB136371
	Zinc	16.7	+/-20	U	40.0	P	07/03/2025	17:32	LB136371