

WATER DIESEL RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE

Lab Name: Alliance

Client: CDM Smith

Lab Code: ACE

SDG No: Q2458

Client Sample ID : PB168697BSD

Datafile: FF016067.D

COMPOUND	SPIKE ADDED ug/L	CONCENTRATION ug/L	LCS/LCSD CONCENTRATION ug/L	% REC	QC LIMITS (%)
DRO	200	0	169	84	78-117

LCS/LCSD % Recovery RPD : 11.9