

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF061325\
 Data File : FF015968.D
 Signal(s) : FID2B.ch
 Acq On : 13 Jun 2025 15:03
 Operator : YP\AJ
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 75 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 5 TRPH STD

Integration File: autoint1.e
 Quant Time: Jun 13 15:22:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF061325.M
 Quant Title :
 QLast Update : Fri Jun 13 15:22:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.089	650939	5.391 ug/ml
Target Compounds			
1) N-OCTANE	2.090	683652	5.388 ug/ml
2) N-DECANE	4.625	707455	5.399 ug/ml
3) N-DODECANE	6.798	719066	5.400 ug/ml
4) N-TETRADECANE	8.628	726616	5.422 ug/ml
5) N-HEXADECANE	10.235	731089	5.413 ug/ml
6) N-OCTADECANE	11.678	752760	5.430 ug/ml
7) N-EICOSANE	12.989	757546	5.425 ug/ml
8) N-DOCOSANE	14.188	735106	5.410 ug/ml
10) N-TETRACOSANE	15.293	734577	5.420 ug/ml
11) N-HEXACOSANE	16.315	725491	5.405 ug/ml
12) N-OCTACOSANE	17.264	716513	5.427 ug/ml
13) N-TRIACONTANE	18.154	732902	5.550 ug/ml
14) N-DOTRIACONTANE	18.987	716819	5.634 ug/ml
15) N-TETRATRIACONTANE	19.772	609289	5.448 ug/ml
16) N-HEXATRIACONTANE	20.512	482765	5.164 ug/ml
17) N-OCTATRIACONTANE	21.272	404248	4.978 ug/ml
18) N-TETRACONTANE	22.225	327064	4.664 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF061325\
Data File : FF015968.D
Signal(s) : FID2B.ch
Acq On : 13 Jun 2025 15:03
Operator : YP\AJ
Sample : 5 TRPH STD
Misc :
ALS Vial : 75 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
5 TRPH STD

Integration File: autoint1.e
Quant Time: Jun 13 15:22:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF061325.M
Quant Title :
QLast Update : Fri Jun 13 15:22:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

