

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG070225\
 Data File : FG016208.D
 Signal(s) : FID1A.ch
 Acq On : 02 Jul 2025 19:06
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Jul 03 01:24:11 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG061325.M
 Quant Title :
 QLast Update : Fri Jun 13 12:34:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.207	5705081	48.636 ug/ml
Target Compounds			
2) N-DECANE	4.676	6254090	49.848 ug/ml
3) N-DODECANE	6.858	6363352	49.868 ug/ml
4) N-TETRADECANE	8.699	6382162	49.454 ug/ml
5) N-HEXADECANE	10.319	6435012	49.302 ug/ml
6) N-OCTADECANE	11.773	6578293	49.120 ug/ml
7) N-EICOSANE	13.093	6626385	48.983 ug/ml
8) N-DOCOSANE	14.300	6451613	48.936 ug/ml
10) N-TETRACOSANE	15.413	6419221	48.730 ug/ml
11) N-HEXACOSANE	16.440	6322288	48.296 ug/ml
12) N-OCTACOSANE	17.397	6167356	47.740 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG070225\
Data File : FG016208.D
Signal(s) : FID1A.ch
Acq On : 02 Jul 2025 19:06
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Jul 03 01:24:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG061325.M
Quant Title :
QLast Update : Fri Jun 13 12:34:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

