

WATER DIESEL RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE

Lab Name: Alliance

Client: CDM Smith

Lab Code: ACE

SDG No: Q2458

Client Sample ID : PB168697BS

Datafile: FF016124.D

COMPOUND	SPIKE ADDED ug/L	CONCENTRATION ug/L	LCS/LCSD CONCENTRATION ug/L	% REC	QC LIMITS (%)
DRO	200	0	194	97	78-117