

WATER GASOLINE RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLIC

Lab Name: Chemtech **Client:** CDM Smith
Lab Code: CHEM **Cas No:** Q2458 **SAS No :** Q2458 **SDG No:** Q2458
Matrix Spike - EPA Sample No : BSF0701W2 **Datafile:** FB031995.D

COMPOUND	SPIKE ADDED ug/L	CONCENTRATION ug/L	LCS/LCSD CONCENTRATION ug/L	% REC	QC LIMITS
GRO	180	0	170	94	50-150

LCS/LCSD % Recovery RPD : 2.9