

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2458
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/07/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030928.D **Time Analyzed:** 19:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.651	9.568	9.768	748.840	712.500	5.1
2,4,5-TP(Silvex)	9.357	9.275	9.475	704.070	712.500	-1.2
2,4-D	8.471	8.386	8.586	678.950	705.000	-3.7
2,4-DB	10.228	10.147	10.347	805.880	712.500	13.1
2,4-DCAA	7.336	7.249	7.449	705.080	750.000	-6.0
DICAMBA	7.526	7.439	7.639	681.450	705.000	-3.3
DICHLORPROP	8.239	8.153	8.353	640.190	705.000	-9.2
Dinoseb	11.444	11.366	11.566	689.280	705.000	-2.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2458
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/07/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030928.D **Time Analyzed:** 19:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.354	10.259	10.459	685.380	712.500	-3.8
2,4,5-TP(Silvex)	9.928	9.833	10.033	679.920	712.500	-4.6
2,4-D	9.023	8.927	9.127	639.930	705.000	-9.2
2,4-DB	10.921	10.827	11.027	670.940	712.500	-5.8
2,4-DCAA	7.766	7.670	7.870	691.990	750.000	-7.7
DICAMBA	7.968	7.872	8.072	671.890	705.000	-4.7
DICHLORPROP	8.687	8.591	8.791	637.410	705.000	-9.6
Dinoseb	11.306	11.211	11.411	648.840	705.000	-8.0