

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2458
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/09/2025
Lab Sample No.: PSTDCCC750 **Data File :** PS030959.D **Time Analyzed:** 11:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.656	9.568	9.768	725.530	712.500	1.8
2,4,5-TP(Silvex)	9.361	9.275	9.475	678.460	712.500	-4.8
2,4-D	8.474	8.386	8.586	643.260	705.000	-8.8
2,4-DB	10.233	10.147	10.347	758.610	712.500	6.5
2,4-DCAA	7.338	7.249	7.449	683.770	750.000	-8.8
DICAMBA	7.528	7.439	7.639	643.540	705.000	-8.7
DICHLORPROP	8.241	8.153	8.353	643.260	705.000	-8.8
Dinoseb	11.451	11.366	11.566	661.940	705.000	-6.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2458
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/09/2025
Lab Sample No.: PSTDCCC750 **Data File :** PS030959.D **Time Analyzed:** 11:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.357	10.259	10.459	629.350	712.500	-11.7
2,4,5-TP(Silvex)	9.931	9.833	10.033	622.060	712.500	-12.7
2,4-D	9.024	8.927	9.127	614.410	705.000	-12.8
2,4-DB	10.925	10.827	11.027	607.740	712.500	-14.7
2,4-DCAA	7.764	7.670	7.870	694.400	750.000	-7.4
DICAMBA	7.967	7.872	8.072	602.990	705.000	-14.5
DICHLORPROP	8.687	8.591	8.791	588.230	705.000	-16.6
Dinoseb	11.311	11.211	11.411	586.380	705.000	-16.8