

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** CAMP02  
**Lab Code:** ACE **SDG NO.:** Q2458  
**GC Column:** RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

**Client Sample No.:** CCAL06 **Date Analyzed:** 07/09/2025  
**Lab Sample No.:** PSTDCCC750 **Data File :** PS030967.D **Time Analyzed:** 14:52

| COMPOUND         | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|------------------|--------|-----------|--------|--------------------|-------------------|-------|
|                  |        | FROM      | TO     |                    |                   |       |
| 2,4,5-T          | 9.653  | 9.568     | 9.768  | 709.820            | 712.500           | -0.4  |
| 2,4,5-TP(Silvex) | 9.359  | 9.275     | 9.475  | 668.070            | 712.500           | -6.2  |
| 2,4-D            | 8.471  | 8.386     | 8.586  | 643.940            | 705.000           | -8.7  |
| 2,4-DB           | 10.231 | 10.147    | 10.347 | 759.090            | 712.500           | 6.5   |
| 2,4-DCAA         | 7.337  | 7.249     | 7.449  | 668.070            | 750.000           | -10.9 |
| DICAMBA          | 7.526  | 7.439     | 7.639  | 648.820            | 705.000           | -8.0  |
| DICHLORPROP      | 8.240  | 8.153     | 8.353  | 607.260            | 705.000           | -13.9 |
| Dinoseb          | 11.447 | 11.366    | 11.566 | 644.130            | 705.000           | -8.6  |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** CAMP02  
**Lab Code:** ACE **SDG NO.:** Q2458  
**GC Column:** RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

**Client Sample No.:** CCAL06 **Date Analyzed:** 07/09/2025  
**Lab Sample No.:** PSTDCCC750 **Data File :** PS030967.D **Time Analyzed:** 14:52

| COMPOUND         | RT     | RT WINDOW<br>FROM TO |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|------------------|--------|----------------------|--------|--------------------|-------------------|-------|
| 2,4,5-T          | 10.361 | 10.259               | 10.459 | 635.110            | 712.500           | -10.9 |
| 2,4,5-TP(Silvex) | 9.934  | 9.833                | 10.033 | 629.760            | 712.500           | -11.6 |
| 2,4-D            | 9.028  | 8.927                | 9.127  | 626.500            | 705.000           | -11.1 |
| 2,4-DB           | 10.929 | 10.827               | 11.027 | 617.890            | 712.500           | -13.3 |
| 2,4-DCAA         | 7.769  | 7.670                | 7.870  | 636.310            | 750.000           | -15.2 |
| DICAMBA          | 7.972  | 7.872                | 8.072  | 613.280            | 705.000           | -13.0 |
| DICHLORPROP      | 8.692  | 8.591                | 8.791  | 588.860            | 705.000           | -16.5 |
| Dinoseb          | 11.313 | 11.211               | 11.411 | 588.950            | 705.000           | -16.5 |