

**SW-846**

Analytical Method: 8151A

Datafile : PS030952.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits	
								Qual	Low	High
PB168717BS (Column 1)	DICAMBA	5	5.10	ug/L	102				67	136
	DICHLORPROP	5	4.90	ug/L	98				88	119
	2,4-D	5	5.20	ug/L	104				83	130
	2,4,5-TP(Silvex)	5	5.30	ug/L	106				78	127
	2,4,5-T	5	5.70	ug/L	114				74	129
	2,4-DB	5	6.00	ug/L	120				53	149
	Dinoseb	5	5.10	ug/L	102				72	131
PB168717BS (Column 2)	DICAMBA	5	4.90	ug/L	98				67	136
	DICHLORPROP	5	4.70	ug/L	94				88	119
	2,4-D	5	4.60	ug/L	92				83	130
	2,4,5-TP(Silvex)	5	5.00	ug/L	100				78	127
	2,4,5-T	5	5.10	ug/L	102				74	129
	2,4-DB	5	4.90	ug/L	98				53	149
	Dinoseb	5	4.60	ug/L	92				72	131