

**SW-846**

Analytical Method: 8151A

Datafile : PS030953.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		
								Qual	Low	High	RPD
PB168717BSD (Column 1)	DICAMBA	5	5.00	ug/L	100	2			67	136	20
	DICHLORPROP	5	4.90	ug/L	98	0			88	119	20
	2,4-D	5	5.90	ug/L	118	13			83	130	20
	2,4,5-TP(Silvex)	5	5.30	ug/L	106	0			78	127	20
	2,4,5-T	5	5.60	ug/L	112	2			74	129	20
	2,4-DB	5	6.00	ug/L	120	0			53	149	20
	Dinoseb	5	5.10	ug/L	102	0			72	131	20
PB168717BSD (Column 2)	DICAMBA	5	4.90	ug/L	98	0			67	136	20
	DICHLORPROP	5	4.70	ug/L	94	0			88	119	20
	2,4-D	5	4.90	ug/L	98	6			83	130	20
	2,4,5-TP(Silvex)	5	5.00	ug/L	100	0			78	127	20
	2,4,5-T	5	5.10	ug/L	102	0			74	129	20
	2,4-DB	5	4.90	ug/L	98	0			53	149	20
	Dinoseb	5	4.70	ug/L	94	2			72	131	20