

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2458 SAS No.: Q2458 SDG NO.: Q2458

Continuing Calib Date: 07/02/2025 Initial Calibration Date(s): 06/11/2025 06/11/2025

Continuing Calib Time: 04:07 Initial Calibration Time(s): 10:40 19:07

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1268-1	(1)	7.60	7.60	7.50	7.70	0.00
Aroclor-1268-2	(2)	7.66	7.67	7.57	7.77	0.01
Aroclor-1268-3	(3)	7.87	7.87	7.77	7.97	0.00
Aroclor-1268-4	(4)	8.16	8.16	8.06	8.26	0.00
Aroclor-1268-5	(5)	8.45	8.46	8.36	8.56	0.01
Tetrachloro-m-xylene		3.68	3.68	3.58	3.78	0.01
Decachlorobiphenyl		8.71	8.71	8.61	8.81	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2458 SAS No.: Q2458 SDG NO.: Q2458

Continuing Calib Date: 07/02/2025 Initial Calibration Date(s): 06/11/2025 06/11/2025

Continuing Calib Time: 04:07 Initial Calibration Time(s): 10:40 19:07

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1268-1	(1)	7.56	7.56	7.46	7.66	0.00
Aroclor-1268-2	(2)	7.62	7.63	7.53	7.73	0.01
Aroclor-1268-3	(3)	7.83	7.83	7.73	7.93	0.00
Aroclor-1268-4	(4)	8.12	8.12	8.02	8.22	0.01
Aroclor-1268-5	(5)	8.40	8.41	8.31	8.51	0.01
Tetrachloro-m-xylene		3.67	3.67	3.57	3.77	0.00
Decachlorobiphenyl		8.66	8.66	8.56	8.76	0.00