

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2458 SAS No.: Q2458 SDG NO.: Q2458

Continuing Calib Date: 07/02/2025 Initial Calibration Date(s): 06/11/2025 06/11/2025

Continuing Calib Time: 11:15 Initial Calibration Time(s): 10:40 19:07

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.76	4.77	4.67	4.87	0.01
Aroclor-1242-2	(2)	4.78	4.78	4.68	4.88	0.00
Aroclor-1242-3	(3)	4.84	4.84	4.74	4.94	0.00
Aroclor-1242-4	(4)	4.96	4.96	4.86	5.06	0.00
Aroclor-1242-5	(5)	5.61	5.61	5.51	5.71	0.00
Tetrachloro-m-xylene		3.68	3.68	3.58	3.78	0.00
Decachlorobiphenyl		8.71	8.71	8.61	8.81	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2458 SAS No.: Q2458 SDG NO.: Q2458

Continuing Calib Date: 07/02/2025 Initial Calibration Date(s): 06/11/2025 06/11/2025

Continuing Calib Time: 11:15 Initial Calibration Time(s): 10:40 19:07

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.75	4.75	4.65	4.85	0.00
Aroclor-1242-2	(2)	4.77	4.77	4.67	4.87	0.00
Aroclor-1242-3	(3)	4.94	4.94	4.84	5.04	0.00
Aroclor-1242-4	(4)	5.03	5.03	4.93	5.13	0.00
Aroclor-1242-5	(5)	5.55	5.55	5.45	5.65	0.00
Tetrachloro-m-xylene		3.67	3.67	3.57	3.77	0.00
Decachlorobiphenyl		8.66	8.66	8.56	8.76	0.00