

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2458 SAS No.: Q2458 SDG NO.: Q2458

Continuing Calib Date: 07/02/2025 Initial Calibration Date(s): 06/11/2025 06/11/2025

Continuing Calib Time: 16:19 Initial Calibration Time(s): 10:40 19:07

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.76	4.77	4.67	4.87	0.01
Aroclor-1016-2	(2)	4.78	4.78	4.68	4.88	0.00
Aroclor-1016-3	(3)	4.84	4.84	4.74	4.94	0.00
Aroclor-1016-4	(4)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-5	(5)	5.22	5.22	5.12	5.32	0.00
Aroclor-1260-1	(1)	6.25	6.26	6.16	6.36	0.01
Aroclor-1260-2	(2)	6.44	6.45	6.35	6.55	0.01
Aroclor-1260-3	(3)	6.81	6.81	6.71	6.91	0.00
Aroclor-1260-4	(4)	7.07	7.07	6.97	7.17	0.00
Aroclor-1260-5	(5)	7.31	7.32	7.22	7.42	0.01
Tetrachloro-m-xylene		3.68	3.68	3.58	3.78	0.00
Decachlorobiphenyl		8.71	8.71	8.61	8.81	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2458 SAS No.: Q2458 SDG NO.: Q2458

Continuing Calib Date: 07/02/2025 Initial Calibration Date(s): 06/11/2025 06/11/2025

Continuing Calib Time: 16:19 Initial Calibration Time(s): 10:40 19:07

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.75	4.75	4.65	4.85	0.00
Aroclor-1016-2	(2)	4.77	4.77	4.67	4.87	0.01
Aroclor-1016-3	(3)	4.94	4.94	4.84	5.04	0.00
Aroclor-1016-4	(4)	4.98	4.99	4.89	5.09	0.01
Aroclor-1016-5	(5)	5.20	5.20	5.10	5.30	0.00
Aroclor-1260-1	(1)	6.22	6.23	6.13	6.33	0.01
Aroclor-1260-2	(2)	6.41	6.42	6.32	6.52	0.01
Aroclor-1260-3	(3)	6.56	6.57	6.47	6.67	0.01
Aroclor-1260-4	(4)	7.04	7.04	6.94	7.14	0.01
Aroclor-1260-5	(5)	7.28	7.28	7.18	7.38	0.00
Tetrachloro-m-xylene		3.67	3.67	3.57	3.77	0.00
Decachlorobiphenyl		8.66	8.66	8.56	8.76	0.00