

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2458 SAS No.: Q2458 SDG NO.: Q2458

Continuing Calib Date: 07/02/2025 Initial Calibration Date(s): 06/11/2025 06/11/2025

Continuing Calib Time: 22:46 Initial Calibration Time(s): 10:40 19:07

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1254-1       | (1) | 5.57       | 5.57      | 5.47      | 5.67 | 0.00       |
| Aroclor-1254-2       | (2) | 5.72       | 5.72      | 5.62      | 5.82 | 0.00       |
| Aroclor-1254-3       | (3) | 6.12       | 6.13      | 6.03      | 6.23 | 0.01       |
| Aroclor-1254-4       | (4) | 6.35       | 6.36      | 6.26      | 6.46 | 0.01       |
| Aroclor-1254-5       | (5) | 6.77       | 6.78      | 6.68      | 6.88 | 0.01       |
| Tetrachloro-m-xylene |     | 3.68       | 3.68      | 3.58      | 3.78 | 0.01       |
| Decachlorobiphenyl   |     | 8.71       | 8.71      | 8.61      | 8.81 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2458 SAS No.: Q2458 SDG NO.: Q2458

Continuing Calib Date: 07/02/2025 Initial Calibration Date(s): 06/11/2025 06/11/2025

Continuing Calib Time: 22:46 Initial Calibration Time(s): 10:40 19:07

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.55       | 5.55      | 5.45      | 5.65 | 0.00       |
| Aroclor-1254-2 (2)   | 5.69       | 5.70      | 5.60      | 5.80 | 0.01       |
| Aroclor-1254-3 (3)   | 6.10       | 6.10      | 6.00      | 6.20 | 0.01       |
| Aroclor-1254-4 (4)   | 6.32       | 6.33      | 6.23      | 6.43 | 0.01       |
| Aroclor-1254-5 (5)   | 6.74       | 6.74      | 6.64      | 6.84 | 0.00       |
| Tetrachloro-m-xylene | 3.67       | 3.67      | 3.57      | 3.77 | 0.00       |
| Decachlorobiphenyl   | 8.66       | 8.66      | 8.56      | 8.76 | 0.00       |