

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2458 SAS No.: Q2458 SDG NO.: Q2458

Continuing Calib Date: 07/02/2025 Initial Calibration Date(s): 07/01/2025 07/01/2025

Continuing Calib Time: 19:36 Initial Calibration Time(s): 14:04 21:30

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|------------------------|-------|------------|
| Aroclor-1254-1       | (1) | 6.49       | 6.49      | 6.39                   | 6.59  | 0.00       |
| Aroclor-1254-2       | (2) | 6.70       | 6.71      | 6.61                   | 6.81  | 0.01       |
| Aroclor-1254-3       | (3) | 7.07       | 7.07      | 6.97                   | 7.17  | 0.00       |
| Aroclor-1254-4       | (4) | 7.35       | 7.35      | 7.25                   | 7.45  | 0.00       |
| Aroclor-1254-5       | (5) | 7.76       | 7.77      | 7.67                   | 7.87  | 0.01       |
| Tetrachloro-m-xylene |     | 4.49       | 4.49      | 4.39                   | 4.59  | 0.00       |
| Decachlorobiphenyl   |     | 10.18      | 10.18     | 10.08                  | 10.28 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2458 SAS No.: Q2458 SDG NO.: Q2458

Continuing Calib Date: 07/02/2025 Initial Calibration Date(s): 07/01/2025 07/01/2025

Continuing Calib Time: 19:36 Initial Calibration Time(s): 14:04 21:30

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1254-1       | (1) | 5.66       | 5.66      | 5.56      | 5.76 | 0.00       |
| Aroclor-1254-2       | (2) | 5.81       | 5.81      | 5.71      | 5.91 | 0.00       |
| Aroclor-1254-3       | (3) | 6.21       | 6.21      | 6.11      | 6.31 | 0.00       |
| Aroclor-1254-4       | (4) | 6.44       | 6.44      | 6.34      | 6.54 | 0.00       |
| Aroclor-1254-5       | (5) | 6.86       | 6.86      | 6.76      | 6.96 | 0.00       |
| Tetrachloro-m-xylene |     | 3.78       | 3.78      | 3.68      | 3.88 | 0.00       |
| Decachlorobiphenyl   |     | 8.79       | 8.79      | 8.69      | 8.89 | 0.00       |