

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2458 SAS No.: Q2458 SDG NO.: Q2458

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/11/2025 06/11/2025

Client Sample No.: CCAL01 Date Analyzed: 07/01/2025

Lab Sample No.: AR1660CCC500 Data File : PO111928.D Time Analyzed: 14:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.762	4.666	4.866	476.070	500.000	-4.8
Aroclor-1016-2	4.781	4.684	4.884	494.350	500.000	-1.1
Aroclor-1016-3	4.838	4.741	4.941	479.610	500.000	-4.1
Aroclor-1016-4	4.958	4.861	5.061	482.220	500.000	-3.6
Aroclor-1016-5	5.215	5.118	5.318	496.660	500.000	-0.7
Aroclor-1260-1	6.254	6.157	6.357	469.840	500.000	-6.0
Aroclor-1260-2	6.444	6.346	6.546	522.650	500.000	4.5
Aroclor-1260-3	6.810	6.713	6.913	549.430	500.000	9.9
Aroclor-1260-4	7.070	6.972	7.172	550.580	500.000	10.1
Aroclor-1260-5	7.313	7.216	7.416	503.270	500.000	0.7
Decachlorobiphenyl	8.709	8.612	8.812	41.680	50.000	-16.6
Tetrachloro-m-xylene	3.675	3.577	3.777	52.290	50.000	4.6

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2458 SAS No.: Q2458 SDG NO.: Q2458

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/11/2025 06/11/2025

Client Sample No.: CCAL01 Date Analyzed: 07/01/2025

Lab Sample No.: AR1660CCC500 Data File : PO111928.D Time Analyzed: 14:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.749	4.651	4.851	504.840	500.000	1.0
Aroclor-1016-2	4.767	4.669	4.869	516.010	500.000	3.2
Aroclor-1016-3	4.942	4.844	5.044	508.300	500.000	1.7
Aroclor-1016-4	4.985	4.887	5.087	505.240	500.000	1.0
Aroclor-1016-5	5.197	5.099	5.299	503.640	500.000	0.7
Aroclor-1260-1	6.227	6.129	6.329	487.640	500.000	-2.5
Aroclor-1260-2	6.415	6.316	6.516	501.580	500.000	0.3
Aroclor-1260-3	6.567	6.469	6.669	473.250	500.000	-5.4
Aroclor-1260-4	7.037	6.939	7.139	460.290	500.000	-7.9
Aroclor-1260-5	7.278	7.181	7.381	459.400	500.000	-8.1
Decachlorobiphenyl	8.658	8.561	8.761	45.950	50.000	-8.1
Tetrachloro-m-xylene	3.672	3.573	3.773	50.710	50.000	1.4