

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2458 SAS No.: Q2458 SDG NO.: Q2458

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/11/2025 06/11/2025

Client Sample No.: CCAL02 Date Analyzed: 07/01/2025

Lab Sample No.: AR1660CCC500 Data File : PO111944.D Time Analyzed: 20:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.762	4.666	4.866	481.590	500.000	-3.7
Aroclor-1016-2	4.781	4.684	4.884	485.830	500.000	-2.8
Aroclor-1016-3	4.838	4.741	4.941	491.600	500.000	-1.7
Aroclor-1016-4	4.957	4.861	5.061	474.670	500.000	-5.1
Aroclor-1016-5	5.215	5.118	5.318	476.220	500.000	-4.8
Aroclor-1260-1	6.253	6.157	6.357	431.680	500.000	-13.7
Aroclor-1260-2	6.443	6.346	6.546	470.870	500.000	-5.8
Aroclor-1260-3	6.809	6.713	6.913	508.620	500.000	1.7
Aroclor-1260-4	7.068	6.972	7.172	514.070	500.000	2.8
Aroclor-1260-5	7.312	7.216	7.416	448.590	500.000	-10.3
Decachlorobiphenyl	8.707	8.612	8.812	43.790	50.000	-12.4
Tetrachloro-m-xylene	3.675	3.577	3.777	51.830	50.000	3.7

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2458 SAS No.: Q2458 SDG NO.: Q2458

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/11/2025 06/11/2025

Client Sample No.: CCAL02 Date Analyzed: 07/01/2025

Lab Sample No.: AR1660CCC500 Data File : PO111944.D Time Analyzed: 20:10

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.748	4.651	4.851	522.410	500.000	4.5
Aroclor-1016-2	4.767	4.669	4.869	521.020	500.000	4.2
Aroclor-1016-3	4.941	4.844	5.044	512.090	500.000	2.4
Aroclor-1016-4	4.984	4.887	5.087	505.020	500.000	1.0
Aroclor-1016-5	5.197	5.099	5.299	493.120	500.000	-1.4
Aroclor-1260-1	6.226	6.129	6.329	466.400	500.000	-6.7
Aroclor-1260-2	6.414	6.316	6.516	481.910	500.000	-3.6
Aroclor-1260-3	6.566	6.469	6.669	456.130	500.000	-8.8
Aroclor-1260-4	7.036	6.939	7.139	443.290	500.000	-11.3
Aroclor-1260-5	7.278	7.181	7.381	442.190	500.000	-11.6
Decachlorobiphenyl	8.656	8.561	8.761	50.680	50.000	1.4
Tetrachloro-m-xylene	3.672	3.573	3.773	51.750	50.000	3.5