

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2458 SAS No.: Q2458 SDG NO.: Q2458

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/11/2025 06/11/2025

Client Sample No.: CCAL06 Date Analyzed: 07/02/2025

Lab Sample No.: AR1660CCC500 Data File : PO111995.D Time Analyzed: 21:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.760	4.666	4.866	471.790	500.000	-5.6
Aroclor-1016-2	4.779	4.684	4.884	484.590	500.000	-3.1
Aroclor-1016-3	4.836	4.741	4.941	471.150	500.000	-5.8
Aroclor-1016-4	4.956	4.861	5.061	476.110	500.000	-4.8
Aroclor-1016-5	5.213	5.118	5.318	489.260	500.000	-2.1
Aroclor-1260-1	6.251	6.157	6.357	464.360	500.000	-7.1
Aroclor-1260-2	6.441	6.346	6.546	498.640	500.000	-0.3
Aroclor-1260-3	6.807	6.713	6.913	496.090	500.000	-0.8
Aroclor-1260-4	7.066	6.972	7.172	479.720	500.000	-4.1
Aroclor-1260-5	7.309	7.216	7.416	482.380	500.000	-3.5
Decachlorobiphenyl	8.705	8.612	8.812	41.570	50.000	-16.9
Tetrachloro-m-xylene	3.673	3.577	3.777	52.800	50.000	5.6

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2458 SAS No.: Q2458 SDG NO.: Q2458

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/11/2025 06/11/2025

Client Sample No.: CCAL06 Date Analyzed: 07/02/2025

Lab Sample No.: AR1660CCC500 Data File : PO111995.D Time Analyzed: 21:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.747	4.651	4.851	515.550	500.000	3.1
Aroclor-1016-2	4.765	4.669	4.869	531.230	500.000	6.2
Aroclor-1016-3	4.940	4.844	5.044	519.890	500.000	4.0
Aroclor-1016-4	4.982	4.887	5.087	517.080	500.000	3.4
Aroclor-1016-5	5.195	5.099	5.299	512.770	500.000	2.6
Aroclor-1260-1	6.224	6.129	6.329	513.060	500.000	2.6
Aroclor-1260-2	6.413	6.316	6.516	534.530	500.000	6.9
Aroclor-1260-3	6.565	6.469	6.669	504.350	500.000	0.9
Aroclor-1260-4	7.035	6.939	7.139	485.350	500.000	-2.9
Aroclor-1260-5	7.277	7.181	7.381	502.790	500.000	0.6
Decachlorobiphenyl	8.656	8.561	8.761	48.520	50.000	-3.0
Tetrachloro-m-xylene	3.670	3.573	3.773	52.260	50.000	4.5