

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2458 SAS No.: Q2458 SDG NO.: Q2458

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 07/01/2025 07/01/2025

Client Sample No.: CCAL08 Date Analyzed: 07/02/2025

Lab Sample No.: AR1660CCC500 Data File : PP073478.D Time Analyzed: 23:58

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|--------|-----------|--------|--------------------|-------------------|-------|
|                      |        | FROM      | TO     |                    |                   |       |
| Aroclor-1016-1       | 5.640  | 5.544     | 5.744  | 476.200            | 500.000           | -4.8  |
| Aroclor-1016-2       | 5.661  | 5.566     | 5.766  | 500.870            | 500.000           | 0.2   |
| Aroclor-1016-3       | 5.723  | 5.629     | 5.829  | 476.170            | 500.000           | -4.8  |
| Aroclor-1016-4       | 5.821  | 5.726     | 5.926  | 488.130            | 500.000           | -2.4  |
| Aroclor-1016-5       | 6.113  | 6.018     | 6.218  | 484.430            | 500.000           | -3.1  |
| Aroclor-1260-1       | 7.230  | 7.134     | 7.334  | 484.390            | 500.000           | -3.1  |
| Aroclor-1260-2       | 7.484  | 7.389     | 7.589  | 474.470            | 500.000           | -5.1  |
| Aroclor-1260-3       | 7.842  | 7.746     | 7.946  | 451.690            | 500.000           | -9.7  |
| Aroclor-1260-4       | 8.065  | 7.971     | 8.171  | 432.530            | 500.000           | -13.5 |
| Aroclor-1260-5       | 8.384  | 8.289     | 8.489  | 440.260            | 500.000           | -11.9 |
| Decachlorobiphenyl   | 10.178 | 10.086    | 10.286 | 45.850             | 50.000            | -8.3  |
| Tetrachloro-m-xylene | 4.489  | 4.393     | 4.593  | 48.710             | 50.000            | -2.6  |

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2458 SAS No.: Q2458 SDG NO.: Q2458

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 07/01/2025 07/01/2025

Client Sample No.: CCAL08 Date Analyzed: 07/02/2025

Lab Sample No.: AR1660CCC500 Data File : PP073478.D Time Analyzed: 23:58

| COMPOUND             | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|----------------------|-------|--------------------|-------------------|------|
| Aroclor-1016-1       | 4.861 | 4.764                | 4.964 | 553.770            | 500.000           | 10.8 |
| Aroclor-1016-2       | 4.879 | 4.782                | 4.982 | 561.550            | 500.000           | 12.3 |
| Aroclor-1016-3       | 5.056 | 4.958                | 5.158 | 576.960            | 500.000           | 15.4 |
| Aroclor-1016-4       | 5.097 | 5.000                | 5.200 | 578.510            | 500.000           | 15.7 |
| Aroclor-1016-5       | 5.311 | 5.214                | 5.414 | 591.760            | 500.000           | 18.4 |
| Aroclor-1260-1       | 6.342 | 6.244                | 6.444 | 550.610            | 500.000           | 10.1 |
| Aroclor-1260-2       | 6.529 | 6.433                | 6.633 | 510.620            | 500.000           | 2.1  |
| Aroclor-1260-3       | 6.682 | 6.585                | 6.785 | 544.940            | 500.000           | 9.0  |
| Aroclor-1260-4       | 7.151 | 7.054                | 7.254 | 545.010            | 500.000           | 9.0  |
| Aroclor-1260-5       | 7.394 | 7.297                | 7.497 | 520.370            | 500.000           | 4.1  |
| Decachlorobiphenyl   | 8.786 | 8.690                | 8.890 | 54.310             | 50.000            | 8.6  |
| Tetrachloro-m-xylene | 3.783 | 3.685                | 3.885 | 53.770             | 50.000            | 7.5  |