

Report of Analysis

Client:	CDM Smith				Date Collected:	06/27/25	
Project: South River W		I Replacement			Date Received:	06/27/25	
Client Sample ID:	: TP-66				SDG No.:	Q2458	
Lab Sample ID:	Q2458-05				Matrix:	SOIL	
Analytical Metho		8082A			% Solid:	88.3 De	canted:
			<u> </u>				
Sample Wt/Vol:	30.06 Unit	U			Final Vol:	10000	uL
Soil Aliquot Vol:		uL			Test:	PCB	
Extraction Type:					Injection Volume :		
GPC Factor :	1.0	PH :					
Prep Method :	SW3541B						
File ID/Qc Batch: Dilution:		Prep Date			Date Analyzed	Prep Batch ID	
PO111942.D	1	07/01/25 08:30			07/01/25 18:47	PB168671	
CAS Number	Parameter	Conc.	Qualifier	MDL		LOQ / CRQL	Units(Dry Weigh
TARGETS							
12674-11-2	Aroclor-1016	4.50	U	4.50		19.2	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60		19.2	ug/kg
11141-16-5	Aroclor-1232	4.20	U	4.20		19.2	ug/kg
53469-21-9	Aroclor-1242	4.50	U	4.50		19.2	ug/kg
12672-29-6	Aroclor-1248	6.70	U	6.70		19.2	ug/kg
11097-69-1	Aroclor-1254	3.60	U	3.60		19.2	ug/kg
37324-23-5	Aroclor-1262	5.70	U	5.70		19.2	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10		19.2	ug/kg
11096-82-5	Aroclor-1260	3.70	U	3.70		19.2	ug/kg
SURROGATES							
877-09-8	Tetrachloro-m-xylene	19.6		32 - 144		98%	SPK: 20
2051-24-3	Decachlorobiphenyl	15.3		32 - 175		76%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates > 25% difference for detected

concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

- J = Estimated Value
- B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration

was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit