

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO061125

Review By	yogesh	Review On	6/11/2025 3:41:50 PM
Supervise By	mohammad	Supervise On	6/13/2025 1:40:28 AM
SubDirectory	PO061125	HP Acquire Method	HP Processing Method PO061125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO111585.D	11 Jun 2025 10:03		YP/AJ	Ok
2	I.BLK	I.BLK	PO111586.D	11 Jun 2025 10:21		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO111587.D	11 Jun 2025 10:40		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO111588.D	11 Jun 2025 10:58		YP/AJ	Ok
5	AR1660ICC500	AR1660ICC500	PO111589.D	11 Jun 2025 11:17		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO111590.D	11 Jun 2025 11:35		YP/AJ	Ok
7	AR1660ICC050	AR1660ICC050	PO111591.D	11 Jun 2025 11:53		YP/AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PO111592.D	11 Jun 2025 12:12		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO111593.D	11 Jun 2025 12:30		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO111594.D	11 Jun 2025 12:48		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO111595.D	11 Jun 2025 13:07		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO111596.D	11 Jun 2025 13:25		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO111597.D	11 Jun 2025 13:44		YP/AJ	Ok
14	AR1242ICC050	AR1242ICC050	PO111598.D	11 Jun 2025 14:02		YP/AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PO111599.D	11 Jun 2025 14:20		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO111600.D	11 Jun 2025 14:39		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO111601.D	11 Jun 2025 15:14		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO111602.D	11 Jun 2025 15:32		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO061125

Review By	yogesh	Review On	6/11/2025 3:41:50 PM
Supervise By	mohammad	Supervise On	6/13/2025 1:40:28 AM
SubDirectory	PO061125	HP Acquire Method	HP Processing Method PO061125
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PO111603.D	11 Jun 2025 15:49		YP/AJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PO111604.D	11 Jun 2025 16:06		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO111605.D	11 Jun 2025 16:25		YP/AJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PO111606.D	11 Jun 2025 16:43		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO111607.D	11 Jun 2025 17:00		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO111608.D	11 Jun 2025 17:18		YP/AJ	Ok
25	AR1262ICC500	AR1262ICC500	PO111609.D	11 Jun 2025 17:36		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO111610.D	11 Jun 2025 17:55		YP/AJ	Ok,M
27	AR1268ICC750	AR1268ICC750	PO111611.D	11 Jun 2025 18:13		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO111612.D	11 Jun 2025 18:31		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO111613.D	11 Jun 2025 18:50		YP/AJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PO111614.D	11 Jun 2025 19:07		YP/AJ	Ok
31	PO061125ICV500	ICVPO061125	PO111615.D	11 Jun 2025 19:25		YP/AJ	Ok
32	AR1242ICV500	ICVPO061125AR1242	PO111616.D	11 Jun 2025 20:02		YP/AJ	Ok
33	AR1248ICV500	ICVPO061125AR1248	PO111617.D	11 Jun 2025 20:39		YP/AJ	Ok
34	AR1254ICV500	ICVPO061125AR1254	PO111618.D	11 Jun 2025 21:16		YP/AJ	Ok
35	AR1268ICV500	ICVPO061125AR1268	PO111619.D	11 Jun 2025 21:52		YP/AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070125

Review By	yogesh	Review On	7/1/2025 12:25:08 PM
Supervise By	mohammad	Supervise On	7/2/2025 7:44:38 AM
SubDirectory	PO070125	HP Acquire Method	HP Processing Method PO061125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO111912.D	01 Jul 2025 09:06		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO111913.D	01 Jul 2025 09:23		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO111914.D	01 Jul 2025 09:42		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO111915.D	01 Jul 2025 10:00		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO111916.D	01 Jul 2025 10:19		YP/AJ	Ok
6	AR1268CCC500	AR1268CCC500	PO111917.D	01 Jul 2025 10:36		YP/AJ	Ok,M
7	I.BLK	I.BLK	PO111918.D	01 Jul 2025 10:54		YP/AJ	Ok
8	Q2448-01	JEI041K-1-1	PO111919.D	01 Jul 2025 11:11	AR1268 Hit	YP/AJ	Ok,M
9	Q2451-03	Y2307-0279-1-1	PO111920.D	01 Jul 2025 11:28		YP/AJ	Ok,M
10	Q2451-04	Y2307-0279-1-2	PO111921.D	01 Jul 2025 11:46		YP/AJ	Ok,M
11	Q2451-05	KQA08OZ-1-1	PO111922.D	01 Jul 2025 12:03		YP/AJ	Ok,M
12	Q2451-06	KQA08OZ-1-2	PO111923.D	01 Jul 2025 12:20		YP/AJ	Ok,M
13	Q2451-07	HIA-925Q-1-1	PO111924.D	01 Jul 2025 12:39		YP/AJ	Ok,M
14	Q2451-08	HIA-925Q-1-2	PO111925.D	01 Jul 2025 12:57		YP/AJ	Ok,M
15	Q2446-03	MR-BUR-LNG-13	PO111926.D	01 Jul 2025 13:15		YP/AJ	Ok,M
16	Q2446-01	MR-BUR-LNG-19	PO111927.D	01 Jul 2025 13:32		YP/AJ	Ok,M
17	AR1660CCC500	AR1660CCC500	PO111928.D	01 Jul 2025 14:42		YP/AJ	Ok
18	AR1242CCC500	AR1242CCC500	PO111929.D	01 Jul 2025 14:59		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070125

Review By	yogesh	Review On	7/1/2025 12:25:08 PM
Supervise By	mohammad	Supervise On	7/2/2025 7:44:38 AM
SubDirectory	PO070125	HP Acquire Method	HP Processing Method PO061125
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PO111930.D	01 Jul 2025 15:18		YP/AJ	Ok
20	AR1254CCC500	AR1254CCC500	PO111931.D	01 Jul 2025 15:35		YP/AJ	Ok
21	AR1268CCC500	AR1268CCC500	PO111932.D	01 Jul 2025 15:52		YP/AJ	Ok,NS
22	I.BLK	I.BLK	PO111933.D	01 Jul 2025 16:10		YP/AJ	Ok
23	PB168671BL	PB168671BL	PO111934.D	01 Jul 2025 16:27		YP/AJ	Ok
24	PB168671BS	PB168671BS	PO111935.D	01 Jul 2025 16:45		YP/AJ	Ok
25	Q2458-01	TP-76	PO111936.D	01 Jul 2025 17:02		YP/AJ	Ok
26	Q2458-02	TP-55	PO111937.D	01 Jul 2025 17:19		YP/AJ	Ok
27	Q2458-02MS	TP-55MS	PO111938.D	01 Jul 2025 17:37		YP/AJ	Ok
28	Q2458-02MSD	TP-55MSD	PO111939.D	01 Jul 2025 17:54		YP/AJ	Ok
29	Q2458-03	TP-68	PO111940.D	01 Jul 2025 18:12		YP/AJ	Ok,NS
30	Q2458-04	TP-67	PO111941.D	01 Jul 2025 18:29		YP/AJ	Ok,NS
31	Q2458-05	TP-66	PO111942.D	01 Jul 2025 18:47		YP/AJ	Ok,NS
32	Q2458-06	TP-60	PO111943.D	01 Jul 2025 19:04		YP/AJ	Ok
33	AR1660CCC500	AR1660CCC500	PO111944.D	01 Jul 2025 20:10		YP/AJ	Ok
34	AR1242CCC500	AR1242CCC500	PO111945.D	01 Jul 2025 21:03	DCB Low in 1st column	YP/AJ	Ok
35	AR1248CCC500	AR1248CCC500	PO111946.D	01 Jul 2025 21:20		YP/AJ	Ok
36	AR1254CCC500	AR1254CCC500	PO111947.D	01 Jul 2025 21:38		YP/AJ	Ok
37	AR1268CCC500	AR1268CCC500	PO111948.D	01 Jul 2025 21:56		YP/AJ	Ok,NS

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070125

Review By	yogesh	Review On	7/1/2025 12:25:08 PM
Supervise By	mohammad	Supervise On	7/2/2025 7:44:38 AM
SubDirectory	PO070125	HP Acquire Method	HP Processing Method PO061125
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	I.BLK	I.BLK	PO111949.D	01 Jul 2025 22:13		YP/AJ	Ok
39	Q2458-07	TP-62	PO111950.D	01 Jul 2025 22:31		YP/AJ	Ok,NS
40	Q2458-08	TP-63	PO111951.D	01 Jul 2025 22:48	AR1254 Hit	YP/AJ	Ok,NS
41	Q2458-09	TP-59	PO111952.D	01 Jul 2025 23:07	need cleanup	YP/AJ	Not Ok
42	Q2459-01	TP-1	PO111953.D	01 Jul 2025 23:24		YP/AJ	Ok,NS
43	Q2462-09	61625-ABC	PO111954.D	01 Jul 2025 23:41		YP/AJ	Ok
44	Q2464-01	OR-3-063025	PO111955.D	02 Jul 2025 00:00		YP/AJ	Ok,NS
45	Q2469-01	WC-1	PO111956.D	02 Jul 2025 00:17	AR1260 Hit	YP/AJ	Ok,NS
46	Q2462-04	60425-AB	PO111957.D	02 Jul 2025 00:36	need cleanup	YP/AJ	Not Ok
47	Q2462-05	50725	PO111958.D	02 Jul 2025 00:53		YP/AJ	Ok,NS
48	Q2462-10	62725	PO111959.D	02 Jul 2025 01:10		YP/AJ	Ok,NS
49	AR1660CCC500	AR1660CCC500	PO111960.D	02 Jul 2025 02:18		YP/AJ	Ok
50	AR1242CCC500	AR1242CCC500	PO111961.D	02 Jul 2025 03:12		YP/AJ	Ok,NS
51	AR1248CCC500	AR1248CCC500	PO111962.D	02 Jul 2025 03:30		YP/AJ	Ok
52	AR1254CCC500	AR1254CCC500	PO111963.D	02 Jul 2025 03:49		YP/AJ	Ok,NS
53	AR1268CCC500	AR1268CCC500	PO111964.D	02 Jul 2025 04:07		YP/AJ	Ok,NS
54	I.BLK	I.BLK	PO111965.D	02 Jul 2025 04:26		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070225

Review By	yogesh	Review On	7/2/2025 11:58:42 AM
Supervise By		Supervise On	
SubDirectory	PO070225	HP Acquire Method	HP Processing Method PO061125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO111966.D	02 Jul 2025 10:38		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO111967.D	02 Jul 2025 10:56		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO111968.D	02 Jul 2025 11:15		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO111969.D	02 Jul 2025 11:32		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO111970.D	02 Jul 2025 11:50		YP/AJ	Ok
6	I.BLK	I.BLK	PO111971.D	02 Jul 2025 12:08		YP/AJ	Ok
7	DDT ANALOGUE	DDT ANALOGUE	PO111972.D	02 Jul 2025 12:26		YP/AJ	Ok
8	PB168671BL	PB168671BL	PO111973.D	02 Jul 2025 12:45		YP/AJ	Ok
9	PB168691BL	PB168691BL	PO111974.D	02 Jul 2025 13:02		YP/AJ	Ok
10	PB168691BS	PB168691BS	PO111975.D	02 Jul 2025 13:19		YP/AJ	Ok
11	Q2458-09	TP-59	PO111976.D	02 Jul 2025 13:37		YP/AJ	Ok
12	Q2467-01	1A-1B-1C-Fire Stopher	PO111977.D	02 Jul 2025 13:54		YP/AJ	Ok
13	Q2462-04	60425-AB	PO111978.D	02 Jul 2025 14:12		YP/AJ	Ok
14	Q2475-01	SOIL-PILE	PO111979.D	02 Jul 2025 15:05		YP/AJ	Ok,NS
15	Q2478-01	WC-1	PO111980.D	02 Jul 2025 15:23		YP/AJ	Ok
16	AR1660CCC500	AR1660CCC500	PO111981.D	02 Jul 2025 16:19		YP/AJ	Ok
17	AR1242CCC500	AR1242CCC500	PO111982.D	02 Jul 2025 16:36		YP/AJ	Ok,NS
18	AR1248CCC500	AR1248CCC500	PO111983.D	02 Jul 2025 16:53		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070225

Review By	yogesh	Review On	7/2/2025 11:58:42 AM
Supervise By		Supervise On	
SubDirectory	PO070225	HP Acquire Method	HP Processing Method PO061125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1254CCC500	AR1254CCC500	PO111984.D	02 Jul 2025 17:11		YP/AJ	Ok
20	I.BLK	I.BLK	PO111985.D	02 Jul 2025 17:28		YP/AJ	Ok
21	PB168704BL	PB168704BL	PO111986.D	02 Jul 2025 17:45		YP/AJ	Ok
22	PB168704BS	PB168704BS	PO111987.D	02 Jul 2025 18:03		YP/AJ	Ok
23	PB168704BSD	PB168704BSD	PO111988.D	02 Jul 2025 18:20		YP/AJ	Ok
24	Q2462-01	40425	PO111989.D	02 Jul 2025 18:39		YP/AJ	Ok
25	Q2463-01	TW-WTS-11	PO111990.D	02 Jul 2025 18:56	Need straight	YP/AJ	Not Ok
26	Q2465-04	62025-ABC	PO111991.D	02 Jul 2025 19:14	TCMX high in 1st column	YP/AJ	Ok,NS
27	Q2481-02	CC0627-CLOXPL	PO111992.D	02 Jul 2025 19:31	Need to run with lower dilution 10X	YP/AJ	Not Ok
28	Q2481-03	CC0625-OXBL	PO111993.D	02 Jul 2025 19:48	Need to run with lower dilution 10X	YP/AJ	Not Ok
29	Q2481-05	CC0625-NL	PO111994.D	02 Jul 2025 20:07	Need to run with lower dilution 10X	YP/AJ	Not Ok
30	AR1660CCC500	AR1660CCC500	PO111995.D	02 Jul 2025 21:16		YP/AJ	Ok
31	AR1242CCC500	AR1242CCC500	PO111996.D	02 Jul 2025 22:11		YP/AJ	Ok
32	AR1248CCC500	AR1248CCC500	PO111997.D	02 Jul 2025 22:28		YP/AJ	Ok
33	AR1254CCC500	AR1254CCC500	PO111998.D	02 Jul 2025 22:46		YP/AJ	Ok
34	I.BLK	I.BLK	PO111999.D	02 Jul 2025 23:05		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070125

Review By	yogesh	Review On	7/1/2025 1:43:36 PM
Supervise By	mohammad	Supervise On	7/2/2025 7:44:45 AM
SubDirectory	PP070125	HP Acquire Method	HP Processing Method PP070125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP073411.D	01 Jul 2025 13:30		YP\AJ	Ok
2	I.BLK	I.BLK	PP073412.D	01 Jul 2025 13:46		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP073413.D	01 Jul 2025 14:04		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP073414.D	01 Jul 2025 14:21		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP073415.D	01 Jul 2025 14:37		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP073416.D	01 Jul 2025 14:54		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PP073417.D	01 Jul 2025 15:10		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP073418.D	01 Jul 2025 15:26		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PP073419.D	01 Jul 2025 15:43		YP\AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PP073420.D	01 Jul 2025 15:59		YP\AJ	Ok
11	AR1242ICC750	AR1242ICC750	PP073421.D	01 Jul 2025 16:16		YP\AJ	Ok
12	AR1242ICC500	AR1242ICC500	PP073422.D	01 Jul 2025 16:33		YP\AJ	Ok
13	AR1242ICC250	AR1242ICC250	PP073423.D	01 Jul 2025 16:49		YP\AJ	Ok
14	AR1242ICC050	AR1242ICC050	PP073424.D	01 Jul 2025 17:05		YP\AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PP073425.D	01 Jul 2025 17:22		YP\AJ	Ok
16	AR1248ICC750	AR1248ICC750	PP073426.D	01 Jul 2025 17:39		YP\AJ	Ok
17	AR1248ICC500	AR1248ICC500	PP073427.D	01 Jul 2025 17:55		YP\AJ	Ok
18	AR1248ICC250	AR1248ICC250	PP073428.D	01 Jul 2025 18:12		YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070125

Review By	yogesh	Review On	7/1/2025 1:43:36 PM
Supervise By	mohammad	Supervise On	7/2/2025 7:44:45 AM
SubDirectory	PP070125	HP Acquire Method	HP Processing Method PP070125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PP073429.D	01 Jul 2025 18:28		YPIAJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PP073430.D	01 Jul 2025 18:45		YPIAJ	Ok
21	AR1254ICC750	AR1254ICC750	PP073431.D	01 Jul 2025 19:01		YPIAJ	Ok
22	AR1254ICC500	AR1254ICC500	PP073432.D	01 Jul 2025 19:18		YPIAJ	Ok
23	AR1254ICC250	AR1254ICC250	PP073433.D	01 Jul 2025 19:35		YPIAJ	Ok
24	AR1254ICC050	AR1254ICC050	PP073434.D	01 Jul 2025 19:51		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP073435.D	01 Jul 2025 20:08		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP073436.D	01 Jul 2025 20:24		YPIAJ	Ok
27	AR1268ICC750	AR1268ICC750	PP073437.D	01 Jul 2025 20:41		YPIAJ	Ok
28	AR1268ICC500	AR1268ICC500	PP073438.D	01 Jul 2025 20:57		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PP073439.D	01 Jul 2025 21:14		YPIAJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PP073440.D	01 Jul 2025 21:30		YPIAJ	Ok,M
31	PP070125ICV500	ICVPP070125	PP073441.D	01 Jul 2025 21:47		YPIAJ	Ok
32	AR1242ICV500	ICVPP070125AR1242	PP073442.D	01 Jul 2025 22:20		YPIAJ	Ok
33	AR1248ICV500	ICVPP070125AR1248	PP073443.D	01 Jul 2025 22:53		YPIAJ	Ok
34	AR1254ICV500	ICVPP070125AR1254	PP073444.D	01 Jul 2025 23:26		YPIAJ	Ok
35	AR1268ICV500	ICVPP070125AR1268	PP073445.D	02 Jul 2025 00:00		YPIAJ	Ok
36	DDT ANALOGUE	DDT ANALOGUE	PP073446.D	02 Jul 2025 07:34		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070225

Review By	yogesh	Review On	7/2/2025 10:58:10 AM
Supervise By		Supervise On	
SubDirectory	PP070225	HP Acquire Method	HP Processing Method PP070125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP073447.D	02 Jul 2025 08:57		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP073448.D	02 Jul 2025 10:07		YP\AJ	Ok
3	AR1242CCC500	AR1242CCC500	PP073449.D	02 Jul 2025 10:27		YP\AJ	Ok
4	AR1248CCC500	AR1248CCC500	PP073450.D	02 Jul 2025 10:44		YP\AJ	Ok
5	AR1254CCC500	AR1254CCC500	PP073451.D	02 Jul 2025 11:00		YP\AJ	Ok
6	I.BLK	I.BLK	PP073452.D	02 Jul 2025 11:17		YP\AJ	Ok
7	DDT ANALOGUE	DDT ANALOGUE	PP073453.D	02 Jul 2025 11:34		YP\AJ	Ok
8	PB168692BL	PB168692BL	PP073454.D	02 Jul 2025 15:14		YP\AJ	Ok
9	PB168692BS	PB168692BS	PP073455.D	02 Jul 2025 15:30		YP\AJ	Ok
10	Q2472-01	PN#1	PP073456.D	02 Jul 2025 15:46		YP\AJ	Ok
11	Q2473-01	PIT#1	PP073457.D	02 Jul 2025 16:03		YP\AJ	Ok,NS
12	Q2473-02	PIT#2	PP073458.D	02 Jul 2025 16:19		YP\AJ	Ok,NS
13	Q2473-03	PIT#3	PP073459.D	02 Jul 2025 16:36		YP\AJ	Ok,NS
14	Q2473-04	PIT#4	PP073460.D	02 Jul 2025 16:53		YP\AJ	Ok,NS
15	Q2473-04MS	PIT#4MS	PP073461.D	02 Jul 2025 17:09		YP\AJ	Ok,NS
16	Q2473-04MSD	PIT#4MSD	PP073462.D	02 Jul 2025 17:25		YP\AJ	Ok,NS
17	AR1660CCC500	AR1660CCC500	PP073463.D	02 Jul 2025 18:47		YP\AJ	Ok,NS
18	AR1242CCC500	AR1242CCC500	PP073464.D	02 Jul 2025 19:04		YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070225

Review By	yogesh	Review On	7/2/2025 10:58:10 AM
Supervise By		Supervise On	
SubDirectory	PP070225	HP Acquire Method	HP Processing Method PP070125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PP073465.D	02 Jul 2025 19:20		YPIAJ	Ok
20	AR1254CCC500	AR1254CCC500	PP073466.D	02 Jul 2025 19:36		YPIAJ	Ok
21	I.BLK	I.BLK	PP073467.D	02 Jul 2025 19:53		YPIAJ	Ok
22	Q2484-01	TP-58	PP073468.D	02 Jul 2025 20:09		YPIAJ	Ok,NS
23	Q2484-02	TP-57	PP073469.D	02 Jul 2025 20:26		YPIAJ	Ok,NS
24	Q2484-03	TP-64	PP073470.D	02 Jul 2025 20:42		YPIAJ	Ok
25	Q2484-04	TP-107	PP073471.D	02 Jul 2025 20:58		YPIAJ	Ok
26	Q2484-05	TP-1006	PP073472.D	02 Jul 2025 21:15		YPIAJ	Ok
27	Q2484-06	TP-104	PP073473.D	02 Jul 2025 21:31		YPIAJ	Ok
28	Q2458-10	FB-06272025	PP073474.D	02 Jul 2025 21:47		YPIAJ	Ok
29	Q2477-01	50728	PP073475.D	02 Jul 2025 22:04		YPIAJ	Ok,NS
30	Q2481-01	CC0627-AL	PP073476.D	02 Jul 2025 22:20		YPIAJ	Ok,NS
31	Q2481-04	CC0627-AOXL	PP073477.D	02 Jul 2025 22:36		YPIAJ	Ok,NS
32	AR1660CCC500	AR1660CCC500	PP073478.D	02 Jul 2025 23:58		YPIAJ	Ok,NS
33	AR1242CCC500	AR1242CCC500	PP073479.D	03 Jul 2025 00:15	failing low in 1st column	YPIAJ	Not Ok
34	AR1248CCC500	AR1248CCC500	PP073480.D	03 Jul 2025 00:31		YPIAJ	Ok
35	AR1254CCC500	AR1254CCC500	PP073481.D	03 Jul 2025 00:47	failing low in 1st column	YPIAJ	Not Ok
36	I.BLK	I.BLK	PP073482.D	03 Jul 2025 01:04		YPIAJ	Ok

M : Manual Integration