

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Group **Contract:** CAMP02
Lab Code: CHEM **SDG NO.:** Q2458
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/17/2025 06/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/01/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089277.D **Time Analyzed:** 16:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.706	6.605	6.805	51.810	50.000	3.6
4,4'-DDE	6.196	6.096	6.296	49.240	50.000	-1.5
4,4'-DDT	7.021	6.921	7.121	44.170	50.000	-11.7
Aldrin	5.272	5.171	5.371	51.690	50.000	3.4
alpha-BHC	4.000	3.899	4.099	52.270	50.000	4.5
alpha-Chlordane	6.028	5.927	6.127	50.240	50.000	0.5
beta-BHC	4.516	4.415	4.615	49.900	50.000	-0.2
Decachlorobiphenyl	9.073	8.972	9.172	48.690	50.000	-2.6
delta-BHC	4.765	4.664	4.864	53.990	50.000	8.0
Dieldrin	6.348	6.247	6.447	50.190	50.000	0.4
Endosulfan I	6.075	5.975	6.175	50.580	50.000	1.2
Endosulfan II	6.787	6.686	6.886	49.450	50.000	-1.1
Endosulfan sulfate	7.150	7.049	7.249	48.310	50.000	-3.4
Endrin	6.574	6.475	6.675	47.220	50.000	-5.6
Endrin aldehyde	6.916	6.815	7.015	48.350	50.000	-3.3
Endrin ketone	7.630	7.530	7.730	49.130	50.000	-1.7
gamma-BHC (Lindane)	4.331	4.230	4.430	51.800	50.000	3.6
gamma-Chlordane	5.947	5.846	6.046	50.960	50.000	1.9
Heptachlor	4.930	4.829	5.029	50.200	50.000	0.4
Heptachlor epoxide	5.691	5.591	5.791	49.950	50.000	-0.1
Methoxychlor	7.493	7.393	7.593	50.180	50.000	0.4
Tetrachloro-m-xylene	3.550	3.450	3.650	51.880	50.000	3.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Group **Contract:** CAMP02
Lab Code: CHEM **SDG NO.:** Q2458
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/17/2025 06/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/01/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089277.D **Time Analyzed:** 16:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.930	5.830	6.030	48.050	50.000	-3.9
4,4'-DDE	5.375	5.275	5.475	50.250	50.000	0.5
4,4'-DDT	6.183	6.084	6.284	44.120	50.000	-11.8
Aldrin	4.369	4.269	4.469	51.130	50.000	2.3
alpha-BHC	3.394	3.293	3.493	51.500	50.000	3.0
alpha-Chlordane	5.189	5.091	5.291	49.750	50.000	-0.5
beta-BHC	4.026	3.926	4.126	49.490	50.000	-1.0
Decachlorobiphenyl	8.072	7.972	8.172	47.700	50.000	-4.6
delta-BHC	4.263	4.162	4.362	50.640	50.000	1.3
Dieldrin	5.511	5.413	5.613	50.100	50.000	0.2
Endosulfan I	5.245	5.147	5.347	49.820	50.000	-0.4
Endosulfan II	6.080	5.981	6.181	49.230	50.000	-1.5
Endosulfan sulfate	6.482	6.383	6.583	48.600	50.000	-2.8
Endrin	5.789	5.689	5.889	50.440	50.000	0.9
Endrin aldehyde	6.258	6.159	6.359	47.730	50.000	-4.5
Endrin ketone	6.991	6.892	7.092	49.210	50.000	-1.6
gamma-BHC (Lindane)	3.728	3.630	3.830	51.080	50.000	2.2
gamma-Chlordane	5.124	5.026	5.226	49.420	50.000	-1.2
Heptachlor	4.083	3.983	4.183	49.000	50.000	-2.0
Heptachlor epoxide	4.871	4.773	4.973	50.730	50.000	1.5
Methoxychlor	6.754	6.655	6.855	51.030	50.000	2.1
Tetrachloro-m-xylene	2.881	2.780	2.980	53.140	50.000	6.3