

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Group Contract: CAMP02

Lab Code: CHEM SDG NO.: Q2458

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 07/01/2025

Lab Sample No.: PSTDCCC050 Data File : PD089289.D Time Analyzed: 19:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.705	6.605	6.805	50.410	50.000	0.8
4,4'-DDE	6.196	6.096	6.296	48.040	50.000	-3.9
4,4'-DDT	7.021	6.921	7.121	42.110	50.000	-15.8
Aldrin	5.271	5.171	5.371	50.960	50.000	1.9
alpha-BHC	3.999	3.899	4.099	51.940	50.000	3.9
alpha-Chlordane	6.027	5.927	6.127	49.270	50.000	-1.5
beta-BHC	4.516	4.415	4.615	49.360	50.000	-1.3
Decachlorobiphenyl	9.072	8.972	9.172	46.970	50.000	-6.1
delta-BHC	4.764	4.664	4.864	53.460	50.000	6.9
Dieldrin	6.347	6.247	6.447	49.220	50.000	-1.6
Endosulfan I	6.075	5.975	6.175	49.550	50.000	-0.9
Endosulfan II	6.786	6.686	6.886	47.770	50.000	-4.5
Endosulfan sulfate	7.149	7.049	7.249	46.540	50.000	-6.9
Endrin	6.575	6.475	6.675	46.290	50.000	-7.4
Endrin aldehyde	6.915	6.815	7.015	47.210	50.000	-5.6
Endrin ketone	7.630	7.530	7.730	47.690	50.000	-4.6
gamma-BHC (Lindane)	4.331	4.230	4.430	51.280	50.000	2.6
gamma-Chlordane	5.946	5.846	6.046	49.990	50.000	0.0
Heptachlor	4.930	4.829	5.029	49.310	50.000	-1.4
Heptachlor epoxide	5.692	5.591	5.791	48.920	50.000	-2.2
Methoxychlor	7.493	7.393	7.593	48.080	50.000	-3.8
Tetrachloro-m-xylene	3.550	3.450	3.650	51.730	50.000	3.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Group **Contract:** CAMP02
Lab Code: CHEM **SDG NO.:** Q2458
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/17/2025 06/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/01/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089289.D **Time Analyzed:** 19:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.930	5.830	6.030	46.490	50.000	-7.0
4,4'-DDE	5.375	5.275	5.475	48.790	50.000	-2.4
4,4'-DDT	6.183	6.084	6.284	41.790	50.000	-16.4
Aldrin	4.369	4.269	4.469	49.870	50.000	-0.3
alpha-BHC	3.393	3.293	3.493	50.900	50.000	1.8
alpha-Chlordane	5.190	5.091	5.291	50.080	50.000	0.2
beta-BHC	4.026	3.926	4.126	48.780	50.000	-2.4
Decachlorobiphenyl	8.072	7.972	8.172	44.010	50.000	-12.0
delta-BHC	4.262	4.162	4.362	49.450	50.000	-1.1
Dieldrin	5.511	5.413	5.613	48.580	50.000	-2.8
Endosulfan I	5.247	5.147	5.347	50.020	50.000	0.0
Endosulfan II	6.080	5.981	6.181	47.250	50.000	-5.5
Endosulfan sulfate	6.482	6.383	6.583	46.320	50.000	-7.4
Endrin	5.789	5.689	5.889	48.800	50.000	-2.4
Endrin aldehyde	6.259	6.159	6.359	45.640	50.000	-8.7
Endrin ketone	6.991	6.892	7.092	46.530	50.000	-6.9
gamma-BHC (Lindane)	3.730	3.630	3.830	49.230	50.000	-1.5
gamma-Chlordane	5.124	5.026	5.226	48.400	50.000	-3.2
Heptachlor	4.083	3.983	4.183	47.790	50.000	-4.4
Heptachlor epoxide	4.871	4.773	4.973	49.490	50.000	-1.0
Methoxychlor	6.753	6.655	6.855	48.100	50.000	-3.8
Tetrachloro-m-xylene	2.881	2.780	2.980	52.340	50.000	4.7