

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Group **Contract:** CAMP02
Lab Code: CHEM **SDG NO.:** Q2458
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/17/2025 06/17/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/01/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089299.D **Time Analyzed:** 22:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.705	6.605	6.805	49.240	50.000	-1.5
4,4'-DDE	6.196	6.096	6.296	47.280	50.000	-5.4
4,4'-DDT	7.021	6.921	7.121	40.210	50.000	-19.6
Aldrin	5.271	5.171	5.371	50.470	50.000	0.9
alpha-BHC	4.000	3.899	4.099	51.690	50.000	3.4
alpha-Chlordane	6.027	5.927	6.127	48.830	50.000	-2.3
beta-BHC	4.516	4.415	4.615	49.400	50.000	-1.2
Decachlorobiphenyl	9.073	8.972	9.172	45.290	50.000	-9.4
delta-BHC	4.764	4.664	4.864	53.160	50.000	6.3
Dieldrin	6.347	6.247	6.447	48.690	50.000	-2.6
Endosulfan I	6.074	5.975	6.175	48.840	50.000	-2.3
Endosulfan II	6.786	6.686	6.886	46.800	50.000	-6.4
Endosulfan sulfate	7.150	7.049	7.249	45.830	50.000	-8.3
Endrin	6.574	6.475	6.675	44.690	50.000	-10.6
Endrin aldehyde	6.915	6.815	7.015	45.270	50.000	-9.5
Endrin ketone	7.630	7.530	7.730	46.190	50.000	-7.6
gamma-BHC (Lindane)	4.330	4.230	4.430	51.150	50.000	2.3
gamma-Chlordane	5.946	5.846	6.046	49.870	50.000	-0.3
Heptachlor	4.930	4.829	5.029	48.000	50.000	-4.0
Heptachlor epoxide	5.690	5.591	5.791	48.490	50.000	-3.0
Methoxychlor	7.493	7.393	7.593	45.780	50.000	-8.4
Tetrachloro-m-xylene	3.550	3.450	3.650	51.450	50.000	2.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Group
Contract: CAMP02
Lab Code: CHEM
SDG NO.: Q2458
GC Column: ZB-MR2
ID: 0.32 (mm)
 Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL04
Date Analyzed: 07/01/2025
Lab Sample No.: PSTDCCC050
Data File : PD089299.D
Time Analyzed: 22:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.929	5.830	6.030	48.080	50.000	-3.8
4,4'-DDE	5.375	5.275	5.475	48.280	50.000	-3.4
4,4'-DDT	6.182	6.084	6.284	40.910	50.000	-18.2
Aldrin	4.368	4.269	4.469	50.130	50.000	0.3
alpha-BHC	3.393	3.293	3.493	50.710	50.000	1.4
alpha-Chlordane	5.189	5.091	5.291	48.800	50.000	-2.4
beta-BHC	4.025	3.926	4.126	48.630	50.000	-2.7
Decachlorobiphenyl	8.071	7.972	8.172	41.860	50.000	-16.3
delta-BHC	4.262	4.162	4.362	49.480	50.000	-1.0
Dieldrin	5.511	5.413	5.613	48.470	50.000	-3.1
Endosulfan I	5.246	5.147	5.347	48.840	50.000	-2.3
Endosulfan II	6.080	5.981	6.181	47.030	50.000	-5.9
Endosulfan sulfate	6.481	6.383	6.583	45.240	50.000	-9.5
Endrin	5.789	5.689	5.889	48.700	50.000	-2.6
Endrin aldehyde	6.258	6.159	6.359	45.360	50.000	-9.3
Endrin ketone	6.990	6.892	7.092	45.230	50.000	-9.5
gamma-BHC (Lindane)	3.729	3.630	3.830	48.650	50.000	-2.7
gamma-Chlordane	5.125	5.026	5.226	49.370	50.000	-1.3
Heptachlor	4.082	3.983	4.183	47.260	50.000	-5.5
Heptachlor epoxide	4.872	4.773	4.973	51.730	50.000	3.5
Methoxychlor	6.753	6.655	6.855	46.030	50.000	-7.9
Tetrachloro-m-xylene	2.881	2.780	2.980	52.440	50.000	4.9