

Report of Analysis

Client:	CDM Smith	Date Collected:	06/27/25
Project:	South River WM Replacement	Date Received:	06/27/25
Client Sample ID:	FB-06272025	SDG No.:	Q2458
Lab Sample ID:	Q2458-10	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	980	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Decanted:	
		Final Vol:	10000
		Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089332.D	1	07/03/25 08:58	07/03/25 14:40	PB168718

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
319-84-6	alpha-BHC	0.0040	U	0.0040	0.051	ug/L
319-85-7	beta-BHC	0.0050	U	0.0050	0.051	ug/L
319-86-8	delta-BHC	0.011	U	0.011	0.051	ug/L
58-89-9	gamma-BHC (Lindane)	0.0038	U	0.0038	0.051	ug/L
76-44-8	Heptachlor	0.0028	U	0.0028	0.051	ug/L
309-00-2	Aldrin	0.0037	U	0.0037	0.051	ug/L
1024-57-3	Heptachlor epoxide	0.0098	U	0.0098	0.051	ug/L
959-98-8	Endosulfan I	0.0032	U	0.0032	0.051	ug/L
60-57-1	Dieldrin	0.0037	U	0.0037	0.051	ug/L
72-55-9	4,4-DDE	0.0038	U	0.0038	0.051	ug/L
72-20-8	Endrin	0.0033	U	0.0033	0.051	ug/L
33213-65-9	Endosulfan II	0.0081	U	0.0081	0.051	ug/L
72-54-8	4,4-DDD	0.0072	U	0.0072	0.051	ug/L
1031-07-8	Endosulfan Sulfate	0.0038	U	0.0038	0.051	ug/L
50-29-3	4,4-DDT	0.0036	U	0.0036	0.051	ug/L
72-43-5	Methoxychlor	0.011	U	0.011	0.051	ug/L
53494-70-5	Endrin ketone	0.0095	U	0.0095	0.051	ug/L
7421-93-4	Endrin aldehyde	0.011	U	0.011	0.051	ug/L
5103-71-9	alpha-Chlordane	0.0036	U	0.0036	0.051	ug/L
5103-74-2	gamma-Chlordane	0.0040	U	0.0040	0.051	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	13.7		57 - 171	69%	SPK: 20
877-09-8	Tetrachloro-m-xylene	21.0		61 - 148	105%	SPK: 20

Report of Analysis

Client:	CDM Smith	Date Collected:	06/27/25
Project:	South River WM Replacement	Date Received:	06/27/25
Client Sample ID:	FB-06272025	SDG No.:	Q2458
Lab Sample ID:	Q2458-10	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	980	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089332.D	1	07/03/25 08:58	07/03/25 14:40	PB168718

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit